

Appendix A

NOTICE OF VIOLATION

Columbus Hospital

License No. 12-00534-04

As a result of the inspection conducted on June 22, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. 10 CFR 35.21 states any licensee authorized to use teletherapy units for treating humans shall perform full calibration measurements at intervals not exceeding one year.

Contrary to the above, a record review and statements made by a licensee representative indicate no full calibration measurement had been performed from March 13, 1978, to June 22, 1981, a period exceeding one year.

This is a Severity Level V violation (Supplement VII).


2. 10 CFR 35.22(b)(5) states that monthly spot-check measurements shall include a determination of the difference between the measured output and the anticipated output expressed as a percentage of the anticipated output (i.e., the value obtained at last full calibration corrected mathematically for physical decay).

Contrary to the above, a review of monthly spot-check records show that no determination of the difference between the measured output and the anticipated output was made from August 1980, to November 1980.

This is a Severity Level V violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

Dated 6/24/81

  
D. J. Greniawski, Chief  
Materials Radiation Protection  
Section No. 2