

Appendix A

NOTICE OF VIOLATION

Dukes Memorial Hospital

License No. 13-18703-01

As a result of the inspection conducted on June 10, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. License Condition 13.A.(1) requires that your nominal 12 millicurie Am-241 source (anatomical maker) be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above, the referenced sealed source has not been leak tested since its receipt in early 1980.

This is a Severity Level V violation (Supplement VII).

2. License Condition 17 requires that all licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated April 9, 1979. This application states that weekly wipe tests (for removable contamination) will be conducted throughout the Nuclear Medicine Department.

Contrary to the above, your weekly wipe test program was not initiated until January 22, 1981.

This is a Severity Level V violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

6/24/81
Date

D. J. Sreniawski
D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2