

NOTICE OF VIOLATION

Midwestern Consulting, Inc.
Ann Arbor, Michigan 48105

License No. 21-26002-01

As a result of the inspection conducted on November 28, 1989, through December 7, 1989, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

License Condition No. 11.A. requires licensed material be used by, or under the supervision and in the physical presence of individuals who have satisfactorily completed the device manufacturer's training program for gauge users and have been designated by the licensee's Radiation Protection Officer. The licensee shall maintain records of the individuals who have been designated as authorized users.


Contrary to the above, on October 4 and 7, and November 7, 1989, licensed material was used by an individual who had not completed the device manufacturer's training program for gauge users. Further, this individual was not under the supervision and physical presence of an individual who completed the training program while using licensed material on the aforementioned dates.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date

12/15/89


for Roy J. Caniano, Chief
Nuclear Materials Section
Section 2