

NOTICE OF VIOLATION

Lansing General Hospital

License No. 21-04073-01

As a result of the inspection conducted on December 20, 1989, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violations were identified:

1. 10 CFR 35.70(a) requires that licensee's survey with a radiation detection survey instrument at the end of each day of use all areas where radiopharmaceuticals are routinely prepared for use or administered.

Contrary to the above, on June 9, October 9 and 23, 1989, direct surveys of use areas had not been conducted as required.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 35.70(e) requires that each licensee survey for removable contamination once each week in all areas where radiopharmaceuticals are routinely prepared for use, administered, or stored.

Contrary to the above, from May 10, 1989 through September 29, 1989, no surveys for removable contamination in all use areas had been conducted.

This is a Severity Level IV violation (Supplement VI).

3. License Condition No. 14 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated December 17, 1987.

The referenced application dated December 17, 1987, states in Section 10.7 that the licensee will implement the model procedure for opening packages that was published in Appendix L to the Regulatory Guide 10.8, Revision 2, which requires, that incoming packages containing radioactive material be surveyed and the results recorded.

Contrary to the above, on July 3 and 10, 1989, surveys of incoming packages containing radioactive materials had not been performed as required.

This is a Severity Level IV violation (Supplement VI).

9001100098 900105  
REG3 LIC30 PDC  
21-04073-01

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

January 5, 1990  
Dated

R. J. Caniano  
R. J. Caniano, Chief  
Nuclear Materials Safety  
Section 2