

D E D / D C B

DEC 22 1989

Harrison County Hospital  
ATTN: Diane Miller  
Assoc. Executive Director  
245 Atwood Street  
Corydon, IN 47112

License No. 13-23555-01

Gentlemen:

This refers to the routine safety inspection conducted by Mr. K. J. Lambert of this office on November 30, 1989, of activities authorized by Byproduct Material License No. 13-23555-01 and to the discussion of our findings with you and Mr. R. Lawalin at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and with the conditions of your license. The inspection consisted of a selective examination of procedures and representative records, observations, independent measurements, and interviews with personnel.

As discussed with you at the conclusion of the inspection certain of your activities appeared to be in violation of NRC requirements. Specifically, you failed to perform weekly wipes of areas of use of radioactive materials between May 19, 1987, and August 11, 1987. When Mr. R. Lawalin assumed the duties of Nuclear Medicine Technologist in August 1987, he identified the failure to perform weekly wipes and took prompt and adequate action to correct this violation and prevent its recurrence.

Because the NRC wants to encourage and support licensee initiative for self-identification and correction of problems, the NRC may exercise discretion in accordance with 10 CFR Part 2, Appendix C, Section V.G. and refrain from issuing a Notice of Violation.

As a result of this inspection, a Notice of Violation is not issued due to your identification and corrective action concerning the violation described above. It is not necessary that you acknowledge receipt of this letter.

In accordance with 10 CFR 2.790 at the Commission's regulations, a copy of this letter will be placed in the NRC Public Document Room.

We will gladly discuss any questions you have concerning this inspection.

Sincerely,

D. J. Sreniawski, Chief  
Nuclear Materials Safety  
Section 1

9001100259 891222  
REG3 LIC30  
13-23555-01 PDC

RIII <sup>yes</sup>  
Lambert/sm  
12/20/89

RIII  
D.J. SRENIAWSKI  
12/20

110  
IE07