

DCM/DCB

Early Detection Center
Predictive and Preventive Medicine

E. G. METROPOULOS, B.S., M.D.
AND ASSOCIATES

11477 E. Twelve Mile Road
Warren, Michigan 48093

(313) 751-0200

December 21, 1989

United States
Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

License No. 21-12972-01


Gentlemen:

This refers to your letter of November 27, 1989 requiring a written response to the identified violations. As you know, I have held an NRC Byproduct Material License for 20 years and I have always had excellent inspections. As noted by Mr. Patterson, I had no byproduct material on-site at the time of the inspection and do not plan to receive any in the immediate future. I do, however, hope to eventually resume my Nuclear Medicine practice and have not cancelled my license.

For each violation:

1. All byproduct material has been removed from the licensed address.
2. If I resume my Nuclear Medicine practice, package monitoring requirements, dose calibrator constancy tests, and area surveys will be performed in accordance with license conditions.
3. Full compliance was achieved at the time of the inspection.

Sincerely,


E. G. Metropoulos, M.D.

EGM/em

9001110099 891221
REG3 LIC30
21-12972-01 FDC

DEC 28 1989

IE 07