Drs. Sachs, Ross and Associates

Diagnostic Radiology, Computed Tomography-CT, CT/MPDI—Spine-Multiplanar Computed Tomography, Mammography, Magnetic Resonance Imaging—MRI, Ultrasonography, Duplex Sonography, Pediatric Radiology, Nuclear Medicine, Angiography, Digital Subtraction Angiography

Ronald J. Ross, M.D. Jay S. Thompson, M.D. Kyung Kim, M.D. Ronald A. Bailey, M.D.

U.S.N.R.C.
Region III
Nuclear Materials Safety Section I
Attn: D.J. Sreniawski, Chief
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Dear Mr. Sreniawski:

This is pursuant to the provisions of 10 CFR 2.201, the following information is submitted as a result of inspection conducted on November 9, 1989 at Community Nuclear Radiology Facility, docket #030-30776, bearing license 34-25962-01.

Item #1 - The survey meters will be calibrated at intervals not to exceed 6 months.

An amendment will be submitted to the Regional NRC Office indicating that this requirement be changed to a one (1) year interval.

Until this amendment has been approved, the survey meters will be calibrated at 6 month intervals.

Thank you for your time and consideration during the routine safety inspection conducted by Mr. K. Lambert and Mr. W. T. King. Should you have any further questions, please feel free to contact me.

Sincerely,

RUINOS

Ronald A. Bailey, M.D. Radiation Safety Officer

9001160042 891228 REG3 LIC30 34-25962-01 PDC

Hillcrest Medical Bidg 6803 Mayfield Road Mayfield Hts., Ohio 44124 Tel. 461-6010 C.T. Center 6801 Mayfield Road Mayfield Hts., Ohio 44124 Tel. 461-5144 Parkway Medical Center 3609 Park East Beachwood, Ohio 44122 Tel. 292-2755

Breast Diagnostic Center 6803 Mayfield Road Mayfield Hts., Ohio 44124 Tel. 461-6010

DEC 1 8 1989