



Sponsored by the Religious Sisters of Mercy founded in 1831 by Catherine McAuley

St. Joseph Mercy Hospital 5301 East Huron River Drive P.O. Box 995 Ann Arbor, Michigan 48106 (313) 572-3456

December 14, 1989

Mr. Roy J. Caniano  
Region III  
Nuclear Regulatory Commission  
799 Roosevelt Road  
Glen Ellyn, IL 60137

RE: Reply to Notice of Violation - Docket #: 030-17041  
License #: 21-00943-04

Dear Mr. Caniano:

In response to the subject matter, we submit the following information:

- 1) License Condition No. 21 has been discussed with all personnel involved and we gained assurance of compliance from all parties.
- 2) To avoid any future violations of the subject condition, any maintenance or repair operations on the teletherapy unit proposed to be performed by our own staff, other than by persons specifically authorized by the Commission, will need the approval of the Radiation Safety Officer or RSO designee, who will verify the nature of work which is not required by the reference condition.
- 3) We have achieved full compliance as of the date of this letter.

If you have any questions, please contact me at (313) 572-2235.

Sincerely,

ChiChi Jacobson, M.S.  
Radiation Safety Officer

Robert L. Ravin  
Administration

/js

9001160381 591214  
REG3 LIC30  
21-00943-04 PDC

110  
IE07  
DEC 22 1989