U.S. NUCLEAR REGULATORY COMMISSION 10 CFR 2.201 SAFETY INSPECTION 2. REGIONAL OFFICE 1. LICENSEE Anlegen Laboratories, Inc P.O Bux 123 30-34 South Main Street U.S. NUCLEAR REGULATORY COMMISSION REGION III 799 ROOSEVELT RD. CLEN ELLYN, IL 60137 , berty, 40 64068 B. DATE OF INSPECTION 4. LICENSE NUMBER(S) 3. DOCKET NUMBER(S) 24-24752-01 030-29357 Licensee: The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commissions (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews, with personnel, and observations by the inspector. The findings as a result of this inspection are as follows: 1. Within the scope of this inspection, no violations were observed. 2. The inspector als., verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time. 3. During this inspection certain of your activities, as checked below, were in violation of NRC requirements. THIS IS A NOTICE OF VIOLATION which is required to be posted in accordance with 10 CFR 19.11. was not properly posted to indicate the presence _. 10 CFR 20.203(b), (c), (d), (e) or 34.42. of a were not properly B. Containers located in ___ labeled to indicate the presence of radioactive material. 10 CFR 20.203(f)(1), or (f)(2). of sealed sources were not performed at the proper License Condition Number frequencies. 10 CFR ___ D. Records of _ were not properly maintained. __ or License Condition Number _ 10 CFR E. Documents were not properly posted or otherwise made available. 10 CFR 19.11. F. Reports or notifications of ___ were not made in accordance or License Condition Number with 10 CFR I hereby state that within 30 days the actions described by me to the inspector will be taken to correct the violations identified in the items checked above This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC. 9001230159 900108 PDC NRC INSPECTOR SIGNATURE - LICENSEE