

Pontiac Osteopathic Hospital

50 N. PERRY ST. PONTIAC, MI 48058
313-338-5000

January 8th, 1990

R.J. Caniano, Chief
U.S. Nuclear Regulatory Commission
Nuclear Materials Safety
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137


Dear Mr. Caniano: Re: License 21-04081-03

In response to your letter of November 22nd, 1989 concerning the inspection of our Nuclear Medicine Department, the following corrective action has been taken:

Daily area surveys of all areas where radiopharmaceuticals are routinely prepared for use or administered are now being performed and recorded using a G M Survey Meter. Prior to the inspection, the technologists were doing daily surveys, however, they were performing a daily wipe test of all the areas which is a more accurate method, however, it was not the exact procedure as directed by 10 CFR 35.70(a).

If there are any questions, please contact our physics consultant, Mr. Ray A. Carlson, M.S. at (313) 662-3197.

Sincerely,



Willie J. Smith
Director - Department of Medical Imaging/Radiology
Pontiac Osteopathic Hospital

WJS/sa
cc. R.A. Carlson, M.S.
Physics Consultant
Encl.



9001300254 900108
REG3 LIC30
21-04081-03 PDC

JAN 12 1990

40
IE07