

November 21, 1989

Bruce S. Mallet, Ph.D., Chief Nuclear Materials Safety Branch United States Nuclear Regulatory Commission Region III 799 Roosevelt Road Glen Ellyn, Illinois 60137 30-11510 21-03194-04

Dear Dr. Mallet:

Attached please find our corrective action steps for the violations identified in the inspection conducted on September 26, 1989.

If you have any further questions or require any further information please do not hesitate to contact me.

Sincerely,

Richard L. Taylor, M.D.

RT:mls

9001300356 891121 REG3 LIC30 21-01394-04 PDC

License #21-03194-01

1. 10 CFR35.22 (a)(1)
 Composition of the Radiation Safety Committee:

The committee composition has been readjusted and the new committee membership as indicated below will begin first quarter (February) 1990 -

Dr. Richard Taylor, Radiology, Radiology Safety Officer Dr. S. Donald Zaentz, Medical Oncology Dr. Mohinder Chadha, Pathology Alice Ford, Nuclear Medicine Technologist Randy Kelley, Administrative Representative Jerry Wolcott, Radiology Manager Diana Mason-Brown, R.N., B.A. Imagene Fischer, R.N.

The recomposition improves the nursing representation. In the past nursing was represented by the quality assurance director who was also a nurse. In the future the representative will be a specific individual separate from quality assurance. The physician representation has been addressed to involve physicians who will be consistently available and interested in committee participation. The administrative representative is the vice president of human resources. A technologist was added to represent another user.

a). Failed to meet quarterly meeting requirements. Cannot deny that there
is no documentation to support meetings during the third quarter 1985 and
1987. Only reasons that can be identified would involve a lack of a
quorum or cancelled due to summer vacations.

To correct this we will, (1) notify committee members either via phone or memo forty-eight hours prior to the meeting, and (2) postpone meetings instead of cancelling. An agenda outline has been developed to facilitate a more efficient and complete meeting.

- b). Patient dose assayed prior to administration I 131 therapy dose.
 We will obtain shielding immediately so this can be accomplished and we will document results.
- c). Survey meters be calibrated at least annually. The Victoreen 492 is not used any more and that is why it is not calibrated. The instrument currently being used for nuclear medicine was out being calibrated and that is why it was not available to be surveyed. We will check to see if there are loan instruments which are available for use while ours is being calibrated.

License #21-03194-04

10 CFR35.615 (d) (4) Record of radiation monitor checks in the entrance of the telepathy room.

Monitor lite is in place and functioning. We have developed a daily log seck list (attached) to document these checks.

RADIATION SAFETY COMMITTEE

11:30 a.m.

November 21, 1989

Nurging Conf. Room

AGENDA

- I. Review of past minutes
- II. Credentials

"Qualifications for new applicants

- Review of individuals currently using NRC licensed radioactive material
- III. Review of new use of by product material (uses). Request for use of NRC licensed radioactive materials within the institution.
 - IV. Report On:

Overexposure

Accident

Spills

Losses/Theft

Unauthorized receipts, uses, transfers, disposals

Mis-administration

Deviation from approved radiation safety practice - (including corrective action of necessary)

Any incidents involving by product material (cause and action taken)

V. Changes in Radiztion Safety Procedures

Policy & Procedure changes

Updates

Revie s

VI. Radiation dose records of all personnel working with by product material occupational radiation exposure.

(Annual review of Radiation Safety Program)

- VII. Quality Assurance Activities
- VIII. Continuing Education

EMMA L. BIXBY HOSPITAL POLICY AND/OR PROCEDURE

TITLE: THERAPY MONITOR & DOOR SAFETY CHECK

POLICY STATEMENT:

NCR, requirement 10 CFR 35.615 (d), requires that a licensee shall maintain a record of safety checks. Paragraph (d)(3) requires a radiation monitor in the entrance of the

Policy #	RAM	5.5	/RAM	6
Effective	Date	2 11	-15-	89
Supercede	5			
Date			-	

teletherapy room be checked with a dedicated check source for proper operation each day before the teletherapy unit is used.

The door to the therapy room will also be check daily.

DAILY CHECKS: (except Saturday, Sunday or Holidays, if the unit is not use

Month:	. 19		
Day:	Performed by:	Day:	Performed by:
1	White Control of the	17.	
2.	-	18.	
3.	The second secon	19.	
4.	-	20.	
5.	Voltage water and a state of the state of th	21.	
6.	processing the control of the contro	22.	
7	***************************************	23.	
8.	******	24.	
9	The state of the s	25.	
10.		26.	
11.		27.	
12.		28.	
13.		29.	
14.		30.	
15.		31.	
16.			

SOURCE NCR, 10 CFR 35.615 (d) (3) /Richard L. Taylor, M.D.

Page 1 of 1

DISTRIBUTION LIST

[7	ACCOUNTING (ACM)
17	ADMINISTRATION (HMM)
11	BUSINESS OFFICE (ACM)
$\mathcal{I}^{\mathcal{I}}$	CENTRAL SUPPLY (CSM)
n	DIETARY (FSM)
7,	Diet Reference Manual
17	DRG/UTILIZATION REVIEW
[7	EDUCATION (EDM)
	ENGINEERING & MAINTENANCE (EMM)
	ENVIRONMENTAL SERVICES (ESM)
	HUMAN RESOURCES (PEM)
	LABORATORY (LBM) Laboratory Reference Manual
<u>1</u> 7	LAUNDRY (LAM)
	MATERIEL SERVICES (MSM)
	MEDICAL RECORDS (MRM) Medical Records Reference Manual
	GENERAL NURSING (GNM)
	Palmer 2 Palmer 3 Palmer 3 Palmer 3 Palmer 4 Palmer 4
	[/ Palmer 3
	[7] Psychiatric Unit Manual
	77 Fediatrics Manual
	// Newborn Nursery Manual
	/7 Obstetrics Manual
	/7 Critical Care Manual
	[7 Operating Room Manual
	/7 Emergency Department Manual
[7	PATIENT REPRESENTATIVE
[]	PAYROLL OFFICE
<u>/</u> 7	PHARMACY (PHM) Pharmacy Reference Manual

	PUBLIC RELATIONS (PRM)
	RADIOLOGY (RAM) Radiology Reference Manual
D	RESPIRATORY THERAPY (RTM) Respiratory Therapy Manual

17 PHYSICAL THERAPY (PTM)

П	SAGE CENTER	R (ALM)
	SOCIAL WORK	(SWM)

[7 SAFETY & SECURITY

_7 V	OLUNTEERS	(VTM)
------	-----------	-------

Reviewed			
Date	Signature		
			
			

A copy of <u>all</u> policies shall be forwarded to Administration.

die File

21-01394-04



November 21, 1989

Bruce S. Mallet, Ph.D., Chief Nuclear Materials Safety Branch United States Nuclear Regulatory Commission Region III 799 Roosevelt Road Glen Ellyn, Illinois 60137

Dear Dr. Mallet:

Attached please find our corrective action steps for the violations identified in the inspection conducted on September 26, 1989.

If you have any further questions or require any further information please do not hesitate to contact me.

Sincerely,

Richard L. Taylor, M.D.

RT:mls

License #21-03194-01

1. 10 CFR35.22 (a)(1)
Composition of the Radiation Safety Committee:

The committee composition has been readjusted and the new committee membership as indicated below will begin first quarter (February) 1990 -

Dr. Richard Taylor, Radiology, Radiology Safety Officer Dr. S. Donald Zaentz, Medical Oncology Dr. Mohinder Chadha, Pathology Alice Ford, Nuclear Medicine Technologist Randy Kelley, Administrative Representative Jerry Wolcott, Radiology Manager Diana Mason-Brown, R.N., B.A. Imogene Fischer, R.N.

The recomposition improves the nursing representation. In the past nursing was represented by the quality assurance director who was also a nurse. In the future the representative will be a specific individual separate from quality assurance. The physician representation has been addressed to involve physicians who will be consistently available and interested in committee participation. The administrative representative is the vice president of human resources. A technologist was added to represent another user.

2. a). Failed to meet quarterly meeting requirements. Cannot deny that there is no documentation to support meetings during the third quarter 1985 and 1987. Only reasons that can be identified would involve a lack of a quorum or cancelled due to summer vacations.

To correct this we will, (1) notify committee members either via phone or memo forty-eight hours prior to the meeting, and (2) postpone meetings instead of cancelling. An agenda outline has been developed to facilitate a more efficient and complete meeting.

- b). Patient dose assayed prior to administration I 131 therapy dose.
 - We will obtain shielding immediately so this can be accomplished and we will document results.
- c). Survey meters be calibrated at least annually. The Victoreen 492 is not used any more and that is why it is not calibrated. The instrument currently being used for nuclear medicine was out being calibrated and that is why it was not available to be surveyed. We will check to see if there are loan instruments which are available for use while ours is being calibrated.

License #21-03194-04

10 CFR35.615 (d) (4) Record of radiation monitor checks in the entrance of the telepathy room.

Monitor lite is in place and functioning. We have developed a daily log check list (attached) to document these checks.

RADIATION SAFETY COMMITTEE

11:30 a.m.

November 21, 1989

Nursing Conf. Room

AGENDA

- I. Review of past minutes
- II. Credentials
 - "Qualifications for new applicants
 - Review of individuals currently using NRC licensed radioactive material
- III. Review of new use of by product material (uses). Request for use of NRC licensed radioactive materials within the institution.
 - IV. Report On:

Overexposure
Accident
Spills
Losses'Theft
Unaut cized receipts, uses, transfers, disposals
Mis-administration
Deviation from approved radiation safety practice - (including corrective action of necessary)
Any incidents involving by product material (cause and action taken)

V. Changes in Radiation Safety Procedures

Policy & Procedure changes Updates Reviews

VI. Radiation dose records of all personnel working with by product material - occupational radiation exposure.

(Annual review of Radiation Safety Program)

- VII. Quality Assurance Activities
- VIII. Continuing Education

EMMA L. BIXBY HOSPITAL POLICY AND/OR PROCEDURE

TITLE: THERAPY MONITOR & DOOR SAFETY CHECK POLICY STATEMENT:

NCR, requirement 10 CFR 35.6'5 (d), requires that a licensee shall maintain a record of safety checks. Paragraph (d)(3) requires a radiation monitor in the entrance of the teletherapy room be checked with a data.

Policy # RAM	5.5/RAM	6
Effective Date	11-15-8	89
Supercedes		T
Date		

teletherapy room be checked with a dedicated check source for propoperation each day before the teletherapy unit is used.

The door to the therapy room will also be check daily.

DAILY C	HECKS:	(except Saturday,	Sunday or Holidays, if	the unit is not
Month:		19		
Day:		Performed by:	Day:	Performed by:
1			17.	
2.			13.	
3.			19.	
4			20.	
5.		-	21.	
6.			22.	
7.			23.	
8.			24.	
9.			25.	
10.			26.	
11.			27.	
12.		***************************************	28.	
13.		***************************************	29.	
14			30.	
15.		-	31.	
16.				

SOURCE NCR, 10 CFR 35.615 (d) (3) /Richard L. Taylor, M.D.

Page 1 of 1