

APPENDIX

NOTICE OF VIOLATION

Olin E. Teague Veterans' Center  
Temple, Texas

Docket: 30-01067/90-01  
License: 42-10739-03

During an NRC inspection conducted on January 26, 1990, violations of NRC requirements were identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1989) (Enforcement Policy), the violations are listed below:

1. 10 CFR 35.51(c) requires that, at the time of survey meter calibration, the apparent exposure rate from a built-in or owner-supplied check source be determined and recorded, and that each survey instrument be checked with the dedicated check source each day of use.

Contrary to the above, as of January 26, 1990, survey meters were routinely being used without first having been checked with a dedicated check source.

This is a Severity Level IV violation. (Supplement VI)

2. 10 CFR 35.59(g) requires that a quarterly inventory be conducted of sealed and brachytherapy sources in the possession of the licensee.

Contrary to the above, the licensee had been conducting semi-annual inventories instead of quarterly inventories since July 1985.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Olin E. Teague Veterans' Center is hereby required to submit to this office, within 30 days of the date of the letter transmitting this Notice, a written statement or explanation in reply, including for each violation: (1) the reason for the violation if admitted, (2) the corrective steps which have been taken and the results achieved, (3) the corrective steps which will be taken to avoid further violations, and (4) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas,  
this 14th day of February 1990