

NOTICE OF VIOLATION

Research Medical Center

License No. 24-18625-01

As a result of the inspection conducted on January 10 and 11, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

License Condition No. 20 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application dated August 1, 1984.

The referenced application dated August 1, 1984, states in Item No. 12 that Clinical, Nursing, Housekeeping, and Security personnel will be required to attend lectures annually for refresher training.

Contrary to the above, licensee training records reviewed during the period November 2, 1984 through December 13, 1989 indicate that security personnel have not attended the annual refresher lectures as required.

This is a Severity Level IV (Supplement VI) violation.

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

February 13, 1990
Dated

R. J. Caniano
R. J. Caniano, Chief
Nuclear Materials Safety
Section 2