

NOTICE OF VIOLATION

Kent Radiology, P.C.
Grand Rapids, MI 49506

License No. 21-24360-01

As a result of the inspection conducted on January 31, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violations were identified:

1. License Condition No. 18 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced documents.

The referenced application dated September 9, 1989, states in Item 10.4 that the licensee will use syringe shields for routine preparation of patient dosages and administration to patients, and wear disposable gloves at all times while handling radioactive materials.

Contrary to the above, since September 1989, syringe shields have not been used for administration to patients nor have disposable gloves been worn while handling radioactive materials.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 35.70(a) requires that licensees survey with a radiation detection survey instrument at the end of each day of use all areas where radiopharmaceuticals are routinely prepared for use or administered.

Contrary to the above, surveys with a radiation detection survey instrument of all areas where radiopharmaceuticals are routinely administered were not performed at the end of each day of use. Specifically, since September 1989, surveys of all areas where radiopharmaceuticals are routinely administered were performed at the beginning of each day of use.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

February 14, 1990

Roy J. Capriano
Roy J. Capriano, Chief
Nuclear Materials Safety Section 2

9002280289 900214
REGS LIC30
21-24360-01 PDC