

APPENDIX

NOTICE OF VIOLATION

Department of the Army
Darnall Army Community Hospital
Fort Hood, Texas

Docket: 30-16084/90-01
License: 42-19113-01

During an NRC inspection conducted on January 25-26, 1990, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1989) (Enforcement Policy), the violation is listed below:

10 CFR 35.51(d) requires that licensees retain a record of each survey instrument calibration for 3 years and that the record include a description of the calibration procedure, the date of the calibration, a description of the source used and the certified exposure rates from the source, the rates indicated by the instrument being calibrated, the correction factors deduced from the calibration data, and the signature of the individual who performed the calibration.

Contrary to the above, as of January 26, 1990, the records of survey instrument calibration for instruments used in nuclear medicine were unavailable for parts of the 3-year retention period.

This is a Severity Level V violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Department of the Army, Darnall Army Community Hospital is hereby required to submit to this office, within 30 days of the date of the letter transmitting this Notice, a written statement or explanation in reply, including for each violation: (1) the reason for the violation if admitted, (2) the corrective steps which have been taken and the results achieved, (3) the corrective steps which will be taken to avoid further violations, and (4) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas,
this 15th day of February 1990

9003020194 900215
REG4 LIC30
42-19113-01 PDC