

F C JEWELL, M D AND ASSOCIATES, P C.
RADIOLOGY

D H BARKER M D
R H GOMLEY M D
G A ROBERTS M D
M MAHMOODZADEGAN, M D
K S BA M D
K K NA AK M D

21510 HARPER AVENUE
BETWEEN B AND D MILE ROADS
ST. CLAIR SHORES MICH 48080
771-6500

June 24, 1981

D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2
United States Nuclear Regulatory
Commission, Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

License No. 21-03312-02

Dear Sir:

I am writing this letter, with regard to your letter dated 5-20-81, regarding the routine safety inspection at Dr. F.C. Jewell and Associates, P. C., 21510 Harper, St. Clair Shores, Michigan 48080. License No. 21-03312-02.

This letter is to further explain the corrective actions, as outlined in the letter by Joseph Mantel, Ph. D. Dr. Mantel's letter is dated June 10, 1981, a copy of which is enclosed.

1. A full calibration chart measured on June 1st, 1981 was received in this office June 1st, 1981. A full calibration chart will always be at the treatment site.
2. The date of calibration and the calibration of the instrument used for full calibration, and spot check measurements was received on the treatment site June 1st, 1981.

8107150192 810710
NMS LIC30
21-03312-02 PDR

JUN 26 1981

License No. 21-03312-02

3. The licensee maintains full documentation of the qualifications of the Physicist, Dr. Joseph Mantel. This information is present on the site, and was present on the site at the date of the routine safety inspection May 12th, 1981.
4. The licensee received on the treatment site June 1st, 1981 a table containing field size/dose rate, and this will be maintained monthly.

I affirm that the above statements are true.

encl: 1

David H. Barker, M.D.

David H. Barker, M. D.
F. C. Jewell, M. D., & Associates, P. C.

Margaret M. Gneweth 6-24-81

MARGARET M. GNEWETH
Notary Public, Wayne County, Michigan
My Commission Expires March 13, 1982