

Appendix A

NOTICE OF VIOLATION

Woodland Hospital

License No. 24-18695-01

As a result of the inspection conducted on June 23, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. License Condition No. 15 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application dated March 19, 1979. Item 10 of the application states the procedures described in Appendix D, Section 2, of the NRC Medical Licensing Guide will be used for calibration of the dose calibrator. Appendix D, Section 2, requires a geometrical variation (at installation), annual accuracy, quarterly linearity, and daily constancy checks of the dose calibrator.

Contrary to the above requirements, full compliance with Appendix D, Section 2, was not met. Specifically, completed records were not available for the geometrical variation, annual accuracy and quarterly linearity tests, and the daily constancy checks were in excess of ± 5 percent from the calculated activity of the cesium-137 standard.

This is a Severity Level V violation (Supplement VII).


2. 10 CFR 35.14(e)(1)(i) requires that sealed sources used as calibration or reference sources be tested for leakage and/or contamination at intervals not to exceed six months. 10 CFR 35.14(e)(2) requires that records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.

Contrary to these requirement, a 213 microcurie cesium-137 sealed source was not leak tested since February 20, 1979.

This is a Severity Level V violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

Dated 7/8/81


D. G. Wiedeman, Acting Chief
Materials Radiation Protection
Section 1

8107150233 810708
NMS LIC30
24-186955-01 PDR