

Muskogee General Hospital

300 Rockefeller Drive, Muskogee, Oklahoma 74401



Phone (918) 682-5501

JOHN O. ROHDE, Director

January 27, 1981

U. S. Nuclear Regulatory Commission
Attn: Mr. Glen Brown, Chief
Fuel Facility and Material
Safety Branch Region IV
611 Ryan Plaza Drive, Suite 1000
Arlington, Texas 76011

RE: Docket #30-11571
License #35-13157-02

Gentlemen:

This refers to your "Notice of Violation" to Muskogee General Hospital Docket #30-11571, License #35-13157-02 dated December 29, 1980. In compliance with your "Notice" the following is submitted:

1. License Condition 14.A requires in part that results of tests for leakage of teletherapy sources shall be kept in units of microcuries.

RESPONSE: In the future wipe test results will be listed as follows:

- A. If the wipe count rate exceeds the background count rate, the results will be listed in microcuries.
- B. If the background count rate exceeds the wipe count rate, the results will be listed as less than .0001 microcuries.

2. 10 CFR 20.407 requires that a personnel monitoring report be submitted to the NRC within the first calendar quarter of 1980 for the total number of individuals for whom personnel monitoring was required or provided during calendar year 1979.

RESPONSE: As of this date, the personnel monitoring reports will be forwarded to the Director of Management & Program Analysis. Tighter auditing controls of record keeping should assure that violations such as this will not occur. Above controls effective as of January 12, 1981.

Should you require information other than presented, please let me know.

Sincerely,

A handwritten signature in cursive script, appearing to read 'Robert Z. Hamm'.

Robert Z. Hamm
Associate Administrator

Sworn and subscribed to before me this 27 day of January 1981

My commission expires 3-12-84

A handwritten signature in cursive script, appearing to read 'Edna Mae Bell'.

Notary Public

8107200019 810206
NMS LIC30
35-13157-01 PDR