

Muskogee General Hospital

300 Rockefeller Drive, Muskogee, Oklahoma 74401



Phone (918) 682-5501

JOHN O. ROHDE, Director

January 27, 1981

U.S. Nuclear Regulatory Commission
Attn: Mr. Glen Brown, Chief
Fuel Facility and Material
Safety Branch Region IV
611 Ryan Plaza Drive, Suite 1000
Arlington, Texas 76011

RE: Docket #30-2922
License #35-13157-01

Gentlemen:

This letter is in response to your "Notice of Violations" to Muskogee General Hospital concerning your December 16-17, 1980 inspection of this facility.

The Victoreen G.M. survey meter #41115 has been sent to the University of Oklahoma, Health Science Center, Department of Radiological Sciences, P. O. Box 26901, Oklahoma City, OK. 73190, license number 35-00526-04 for calibration. The Victoreen ionization survey meter, model 740F, will be sent to the same company for calibration when the G.M. survey meter is returned.

A Record Keeping Auditing Component has been added to our quality control program. This Auditing component should assure that: (1) routine dose calibrator checks that are required in our license, (2) annual survey meter calibration by a licensed service, and (3) other required record keeping items such as those listed as violations in your letter, will be performed with the required frequency.

We believe these actions appropriately correct the violation items you found. It should also greatly increase the probability that full compliance with all license regulations will be achieved in the future. Full compliance with these actions are expected to be implemented no later than February 28, 1981.

This letter of compliance was written and required conditions met by Mike Morris, our Radiation Physicist. Should you require other than presented, please let me know.

Sincerely,

A handwritten signature in cursive script, appearing to read 'Robert Z. Hamm'.

Robert Z. Hamm
Associate Administrator

Sworn and subscribed to before me this 28 day of January 1981.

My Commission expires 3-12-84.

A handwritten signature in cursive script, appearing to read 'E. W. ...'.

Notary Public

8107200026 810206
NMS LIC30
35-13157-01 PDR