

IMMEDIATE ACTION LETTER

DCS

Lic. File

JUL 14 1981



Dexter Memorial Hospital
ATTN: Mr. Melvin Walker
Administrator
One Mile Road
Dexter, MO 63841

License No. 24-18627-01

Gentlemen:

This is in reference to the telephone conversation between you and Dr. W. J. Adam of this office on July 13, 1981.

It is our understanding that on or before July 16, 1981, you will perform a test to confirm the accuracy of your dose calibrator to within $\pm 5\%$ of the stated activity on a precalibrated dose of technetium-99m. Moreover, you will have your dose calibrator tested for linearity and accuracy within 20 days of the date of this notice by an individual authorized by the NRC to perform such tests.

If our understanding in this matter is not correct, please notify us immediately.

Sincerely,

James G. Keppler
Director

cc: DMB/Document Control Desk (RIDS)

*IE07
s
1/0*

*Conferred in by
Dudley Thompson by phone
on 7/14. aRL.*

RIII
J.M.
Madera/np
7/13/81

RIII
wj
Adam
7/13

RIII
BSU
Greger

RIII
BSU
Norelius
7/13/81

RIII
d
for Streeter
7/14

RIII
d
Davis
7/14

RIII
d
for Keppler
7/14

IMMEDIATE ACTION LETTER