

APPENDIX A

NOTICE OF VIOLATION

University City Science Center
Philadelphia, Pennsylvania 19104
License No. 37-17452-01

Docket No. 30-12779

As a result of the inspection conducted on June 4, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

- A. 10 CFR 20.401(b) requires that you maintain records showing the results of surveys as may be necessary for you to comply with all sections of Part 20.
1. Contrary to this requirement, as of the day of the inspection, June 4, 1981, you failed to maintain records as were necessary to assure compliance with 10 CFR 20.201(b), "Surveys". Specifically, you failed to maintain records of the contamination surveys that you performed within your radioisotope laboratories.
 2. Contrary to this requirement, as of the day of the inspection, June 4, 1981, you failed to maintain records as were necessary to assure compliance with 10 CFR 20.303, "Disposal by release into sanitary sewerage system." Specifically, you failed to maintain records of the disposals of licensed materials that you have made into the sanitary sewerage system.

These are Severity Level VI Violations. (Supplement VII)

Pursuant to the provisions of 10 CFR 2.201, University City Science Center is hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation.

Dated

July 7, 1981

John D. Kinneman
John D. Kinneman, Chief, Materials
Radiological Protection Section,
Technical Inspection Branch