

RADIOLOGY DEPARTMENT

NORMAND E. GIRARD
President

SOMERVILLE HOSPITAL

230 HIGHLAND AVENUE
SOMERVILLE, MASS. 02143

Radiologists

LAWRENCE H. ANDRESON, M.D.
JAMES A. PETERSEN, M.D.
JOHN T. POLLARD, JR., M.D.
VINCENT R. SITES, M.D.

License No. 14009-01

15 June 1981

John D. Kinneman, Chief
Materials Radiological Protection
Section
U.S.N.R.C.
Region 1
631 Park Ave.
King of Prussia, Pennsylvania 19406


Dear Sir:

Re: Inspection 81-01

In reference to your correspondence of 1 June 1981 and the infraction noted therein, the following steps have been taken:

the cesium-137 source has been tested and found to be in excellent condition. This test was performed on 15 April 1981 and was subsequently documented. This test will be repeated at six month intervals with the test results documented to achieve compliance with the regulations.

Sincerely,



Justin J. Dinn, R.T.
Chief Radiologic Technologist