

C-E Power Systems  
Combustion Engineering, Inc.  
1000 Prospect Hill Road  
Windsor, Connecticut 06095



Tel. 203/688-1911  
Telex. 9-9297

07001100A05M  
CLOSED 3/20/81

70-1100

PDR

Return to  
D. Cramer  
396-55



February 2, 1981

U. S. Nuclear Regulatory Commission  
Division of Safeguards  
7915 Eastern Avenue-Wilste Building  
Mail Stop 881-SS  
Silver Springs, Maryland 20910

Applicant.....
Check No. <i>070389</i> .....
Amount/Fee Category <i>150.10</i> .....
Type of Fee <i>add'l</i> .....
Date Check Rec'd. <i>6/10/81</i> .....
Received By... <i>Wm Turner</i> .....

Attention: Mr. James G. Partlow, Chief  
Material Control Licensing Branch

Subject: SNM-1067 - Docket No. 70-1100  
Safeguards Amendment MPP-4

Gentlemen:

Amendment MPP-4 to SNM-1067 sets forth the various safeguards amendments assigned to SNM-1067. Included in this amendment are license conditions which were issued after your review and approval of our submittal required under 70 CFR 70.57 (Measurement Control Plan).

In our February 15, 1979 submittal of the above mentioned plan, it was stated in Attachment 3 - Standards Measurements - Item 2A that one standard would be included with each group of gravimetric analyses run. License Condition 4.7 to the subject amendment requires that a minimum of two standards of measurements per week for each measurement system.

The decision to stipulate one standard measurement run with each group of samples run for gravimetric analyses in the above mentioned Measurement Control Plan was based on technological justification and resources. One standard run with each group experiences the same condition(s) that are in the furnace at the time of the run. Standard results can be reviewed immediately to determine if it is within the acceptable limits. Historical operational data further provides sufficient information for use in our overall statistical program. The furnace used in the gravimetric analysis procedure has capacity for only twelve samples. Accordingly, we attempt to schedule eleven samples and one standard for each group of analyses run thus providing the most economical use of our personnel and equipment.

19228

8107280054 810202  
PDR ADCK 07001100  
C PDR


File 281

Combustion Engineering, Inc., would request, therefore, that License Condition 4.7 in Amendment MPP-4 to License SNM-1067 be revised to authorize the requirement that one standard measurement be included with each group of samples that are run for uranium percentage determination by the gravimetric procedure. ✓

The necessary revisions to our Measurement Control Plan will be submitted in the event the above mentioned revision is granted.

Very truly yours,

COMBUSTION ENGINEERING, INC.



H. V. Lichtenberger  
Vice President-Nuclear Fuel  
Nuclear Power Systems Division

HVL:TBB:sas

William O. Miller  
License Fee Management Branch  
Office of Administration

MC # \_\_\_\_\_  
SG # \_\_\_\_\_  
MCL# 07001100A05M  
MSL# \_\_\_\_\_

MATERIALS SAFEGUARDS AMENDMENT CLASSIFICATION

Applicant: C E - Windsor  
License No: SNM-1067 Fee Category: 1B  
Application Dated: 2/2/81 Received: 2/9/81  
Applicant's Fee Classification: \_\_\_\_\_

The above application for amendment has been reviewed by the NMSS Division of Safeguards in accordance with Section 170.31 of Part 170 and is classified as follows:

1. Safeguard amendments to licenses in fee Categories 1A through 1H
  - (a) \_\_\_\_\_ Major Safeguards
  - (b) \_\_\_\_\_ Minor Safeguards
  - (c) \_\_\_\_\_ Safeguards (Categories 1D through 1G only)
  - (d)  Administrative

2. Justification for reclassification: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

3. The application was filed (a) \_\_\_\_\_ pursuant to written MRC request and the amendment is being issued for the convenience of the Commission, or (b) \_\_\_\_\_ Other (State reason): \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Signature: *Pa. Knight*  
Division of Safeguards, NMSS  
Date: 2/11/81

*Standard Measurements*

APR 24 1981

DOCKET NO. 70-1100

C-E Power Systems  
Combustion Engineering, Inc.  
ATTN: Mr. H. V. Lichtenberger  
Vice President - Nuclear Fuel  
Nuclear Power Systems Division  
1000 Prospect Hill Road  
Windsor, Connecticut 06095

Gentlemen:

Enclosed is a copy of our February 18, 1981 letter, requesting a \$150 fee for a safeguards amendment to License SNM-1067.

As of this date we have not received the fee. Please remit within 15 days from the date of this letter. Payment should be made to the U. S. Nuclear Regulatory Commission and mailed to my attention.

Sincerely,

Original Signed By  
J. D. Weiss

Douglas Weiss  
License Fee Management Branch  
Silver Spring Office  
Office of Administration

Enclosure:  
Letter Dated 2/18/81

DISTRIBUTION:

Docket File  
PDR  
Matls. License Fee File  
LFMB R/F (SS)  
LFMB R/F (Beth)  
ASCabell, LFMB

19223

DUPE: 8105050106

OFFICE	LFMB/ADM						
SURNAME	DWeiss:rej						
DATE	4/24/81						

FEB 18 1981

DOCKET NO. 70-1100

C-E Power Systems  
Combustion Engineering, Inc.  
ATTN: Mr. H. V. Lichtenberger  
Vice President-Nuclear Fuel  
Nuclear Power Systems Division  
1000 Prospect Hill Road  
Windsor, Connecticut 06095

Gentlemen:

This refers to your letter dated February 2, 1981, for an amendment to safeguards Amendment MPP-4 of License SNM-1067.

An administrative amendment fee of \$150 is required as specified in Section 170.31 (1B) of 10 CFR 170, copy enclosed. Payment should be made to the U.S. Nuclear Regulatory Commission and mailed to my attention.

Sincerely,

Original Signed By  
J. D. Weiss

Douglas Weiss  
License Fee Management Branch  
Silver Spring Office  
Office of Administration

Enclosure:  
10 CFR 170

DISTRIBUTION:

✓ Docket File  
PDR  
LFMB R/F (SS)  
LFMB R/F (Beth)  
ASCabell, LFMB

19223

DUPE: 8103060155

OFFICE	LFMB:ADM						
SURNAME	DWeiss:rej						
DATE	2/ /81						