



P. O. Box 816 -- Phone (405) 229-0440
HEALDTON, OKLAHOMA 73438

JUNE 22, 1981

LICENSE NO. 35-19260-01

GENTLEMEN:

THE FOLLOWING CORRECTIVE STEPS HAVE BEEN TAKEN TO INSURE THAT FURTHER VIOLATIONS WILL NOT OCCUR

1. LICENSE CONDITION #12 -- A. RADIATION SCHOOL WILL BE SET UPON 6 MOS. INTERVELS TO INSURE THAT NEW INSPECTORS WILL KNOW RADIATION SAFETY
2. LICENSE CONSITION #17 -- A REMINDER OF NEEDED CALIBRATION OF SURVEY INSTRUMENT WILL BE SENT BY N.S.S.I., 3 WEEKS PRIOR TO TO CALIBRATION DUE DATE. QUARTERLY INVENTORY IS BEING KEPT IN DAILY SURVEY LOG BOOK.
3. LICENSE CONDITION #13 -- LEAK TEST WILL BE MAILED FROM N.S.S.I. HOUSTON & RETURNED FOR TESTS EVERY 6 MOS., RESULTS WILL BE ON FILE
4. LICENSE CONDITION #15 -- A INVENTORY RECORD OF THE SEALED SOURCE, TYPE OF BY PRODUCT MATERIAL & LOCATION WILL BE KEPT ON FILE.
- 5 & 6 A BULLETIN BOARD HAS BEEN INSTALLED IN THE INSPECTION ROOM; THE FOLLOWING ARE NOW POSTED -- MATERIAL LICENSE
 1. #35-19260-01
 2. NCR - 3
 3. 10 CFR PART 19
 4. 10 CFR PART 20

A FILE HAS BEEN SET UP CONTAINING THE FOLLOWING:

1. LEAK TEST REPORTS
2. PERSONNEL MONITORING REPORTS
3. SURVEY INSTRUMENT CALIBRATION REPORT
4. RADIATION SAFETY MANUAL
5. NRC POLICY STATEMENTS

I HEREBY SWEAR THAT ALL ABOVE STEPS HAVE BEEN IMPLEMENTED.

JESS PARNELL
PLANT MANAGER

My commission expires 4/30/85

8107280528 810718

NMS LIC30

35-19260-01

PDR

Notary Public