**Veterans
Administration**

JUN 19 1981

Mike Shopen
Technical Inspection Branch
Nuclear Regulatory Commission
Region IV
611 Ryan Plaza Drive, Suite 1000
Arlington, Texas 76011




Dear Mr. Shopen,

In response to our telephone conversation of June 8, 1981, the following is for your information.

The first violation cited in the report of the NRC inspection of April 24, 1981 dealt with calibration of survey meters. The second violation dealt with surveys and monitoring records of the Gamma cell. In addition to the previously stated methods for assuring that these violations will not re-occur, (letter dated May 22, 1981 to Glen D. Brown, Chief), both have been added to newly established agenda protocols of the Radiation Safety Committee. At appropriate intervals, the Radiation Safety Committee will audit the Radiation Safety Technician's survey meter calibrations and Gamma cell monitoring.

I certify that all information contained in this letter is true and correct to the best of my knowledge and belief.

Sincerely,


MANSELL G. PIPER
Medical Center Director

cc: Glen D. Brown, Chief
Technical Inspection Branch

In Reply Refer To:

8107280540 810718
NMS LIC30
05-01401-06 PDR