



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION I  
631 PARK AVENUE  
KING OF PRUSSIA, PENNSYLVANIA 19406

Docket No. 30-14664

15 JUL 1981

Southwestern Health Center  
ATTN: Mrs. Mary Katherine Petri  
Administrator  
500 Lewis Run Road  
Pittsburgh, Pennsylvania 15236

Gentlemen:

Subject: Inspection 81-01

This refers to the routine safety inspection conducted by Mr. F. M. Costello of this office on June 19, 1981 of activities authorized by NRC License No. 37-18201-01 and to the discussions of our findings held by Mr. Costello with Dr. E. J. Pavsek of your staff at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspector, and observations by the inspector.

Based on the results of this inspection, it appears that certain of your activities were not conducted in full compliance with NRC requirements, as set forth in the Notice of Violation, enclosed herewith as Appendix A. These items of non-compliance have been categorized into the levels described in the Federal Register Notice (45 FR 66754) dated October 7, 1980. You are required to respond to this letter and in preparing your response, you should follow the instructions in Appendix A.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter and your reply will be placed in the Public Document Room.

Should you have any questions concerning this inspection, we will be pleased to discuss them with you.

Sincerely,

*Francis M. Costello*  
for John D. Kinneman, Chief, Materials  
Radiological Protection Section,  
Technical Inspection Branch

Enclosure: Appendix A, Notice of Violation

1607  
S  
1/1

9108030182 810715  
NMS LICCO  
37-18201-01 PDR