

GEI

GROUND ENGINEERING, INC.  
Geotechnical Engineers

12125 Woodcrest Executive Drive, Suite 165

St. Louis, Missouri 63141

Telephone (314) 576-1172

December 5, 1989

Mr. Roy J. Caniano, Chief  
Nuclear Materials Safety  
United States Nuclear Regulatory  
Commission - Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Docket No. 030-18234  
License No. 24-20311-01

Gentlemen:


This refers to your letter dated September 19, 1989 regarding the routine safety inspection performed in our office on September 8, 1989 by Mr. Gary L. Shear. The letter was received in our office on December 1, 1989.

In response to License Condition No. 13, Items 7.A. and 7.B., relating to Radiation Leak Tests: Tests will be taken concurrently with the six-month inventory. If a gauge has been inactive during the period between inventories this will be noted on the inventory record.

In response to 10 CFR 20.401(a) relating to personnel monitoring dosimeter records. All records will be kept filed for review during the six-month inventory. Names, Social Security Numbers and Birth Dates of pertinent personnel have been provided to our dosimeter service. Our dosimeter service is Tech/Ops Landauer, Inc., Glenwood, Ill.

A copy of your September 19th letter and Notice Of Violation are attached. Please feel free to call us if you have any further questions.

Sincerely,



Mark Henson  
Ground Engineering, Inc.

8912270265 891205  
REG3 LIC30  
24-20311-01 PDC

110  
IE07

DEC 11 1989