

EDWARD E. ROY, MHA  
ADMINISTRADOR



DC 5

LIGA PUERTORRIQUEÑA  
CONTRA EL CÁNCER

HOSPITAL ONCOLOGICO  
I. GONZALEZ MARTINEZ

CENTRO MEDICO  
RIO PIEDRAS, PUERTO RICO 00935  
TELEFONO 763-4149  
APARTADO 1811  
HATO REY, PUERTO RICO 00919  
TELEFONO 763-4707

April 26, 1983

Director  
Office of Inspection and Enforcement  
U. S. Nuclear Regulatory Commission  
Washington, D. C. 20555

52-13471-00


Dear Sir:

Reference our letter to you dated April 19, 1983, which contains our reply to the violation stated in Mr. James P. O'Reilly's letter dated March 23, 1983.

Attached please find a copy of a Certificate from the Therapy Services Incorporated, which verifies that the maintenance operations that were performed by the Oncologic Hospital personnel were done such that no unusual significant adverse safety will result.

I hope this meet with your satisfaction. If there is any question concerning to this communication, please contact me at the Area Code 809 - 765-2382.

Respectfully,



Edward E. Roy  
Administrator

EER/nim  
Enclosure

cc: Mr. James P. O'Reilly  
U. S. Nuclear Regulatory Commission

President  
Radiation Safety Committee

Mario E. Rosa García, M.D.  
Medical Director

Antonio Bosch, M.D.  
Department of Radiotherapy

Mr. Jaime Figueroa  
Bio Medical Electronics Department

IE14

8305090341 830426  
NMS LIC30  
52-13471-02 PDR

Therapy Services Incorporated  
Five Year Inspection Certificate

This is to certify that the A.E.C.L. Cobalt-60  
teletherapy unit, Model No. THERATRON 80 Serial No. 169  
located at GE. MARTINEZ ONCOLOGIC HOSPITAL,  
HATO REY, PUERTO RICO was inspected and serviced on  
APRIL 25, 1983 by LESLIE F. FORREST to  
assure the proper function of the source drive mechanism as  
authorized by Maryland Radioactive Materials License Number  
MD-21-009-01.

Signed Leslie F. Forrest Date 4-25-83

Parts Used 1- AIRLINE FILTER  
1- LIGHT CONDUCTOR TAPE  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Recommendations for future service \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

I further certify, that Therapy Services Incorporated, Source  
Drive Mechanism Servicing Procedure 1001 or 1002 was followed.

Signed Leslie F. Forrest Date 4-25-83

The maintenance operations that were performed by the Oncologic  
Hospital personnel were done such that no unusual significant  
adverse safety will result.

Signed Leslie F. Forrest Date 4-25-83