

APPENDIX A

NOTICE OF VIOLATION

Louisiana Power & Light Company  
Waterford Steam Electric Station

Docket: 50-382  
License: CPPR-103

Based on the results of an NRC investigation conducted during the period of March 29, through May 16, 1982, and in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C), 47 FR 9987, dated March 9, 1982, the following violation was identified:

10 CFR Part 50, Appendix B, Criterion V, states, in part: Activities affecting quality shall be prescribed by documented instructions, procedures, or drawings, of a type appropriate to the circumstances and shall be accomplished in accordance with these instructions, procedures, or drawings.

Administrative Procedure MI-1-007, Revision 0, paragraph 5.4, Metrology Lab Qualification, requires specific qualification activities to be completed, including a final signature on a qualification card by the Metrology Supervisor-Nuclear, in order for an individual to be qualified to perform calibration activities in the Metrology Lab.

Contrary to the above, an individual in the Louisiana Power & Light Company Metrology Lab performed equipment calibration without being qualified over a period of about 18 months from approximately October 1980 through April 1982.

This is a Severity Level IV Violation. (Supplement (II) (382/8209-01)

Pursuant to the provisions of 10 CFR 2.201, Louisiana Power & Light Company is hereby required to submit to this office, within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated:     MAY 4 1983