

APPENDIX A

NOTICE OF VIOLATION

Milton J. Lilly, Jr., M.D.

License No. 47-02611-03

As a result of the inspection conducted on July 5, 1983 and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified.

License Condition 13 requires that each sealed source containing licensed material be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above, the licensee's strontium - 90 sealed source has not been tested for leakage and/or contamination since 1980.

This is a Severity Level IV Violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violation; (2) the reasons for the violation if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: JUL 15 1983