U.S. NUCLEAR REGULATORY COMMISSION

REGION III

V.A. Edward Hines Jr., Medical Center ATTN: Joan Cummings, M.D. Director P.O. Box 5000 Hines, IL 60141

Report No. 030-01391/94001(DRSS)

License No. 12-01087-07

Category G1

Priority 1

License No. 12-01087-07

Docket No. 030-01391

Licensee: V.A. Edward Hines Jr., Medical Center Hines, IL 60141

Inspection Dates: April 13, 1993

Inspector:

. 1.

Michael F. Weber Materials Inspector

Approved By:

John D. Jones, Acting Chief Nuclear Materials Inspection Section 2 4/15/34 Date

04/25/94

Date

Inspection Summary

<u>Inspection on April 13, 1994 (Report No. 030-01391/94001(DRSS))</u> <u>Areas Inspected</u>: This was a special, announced safety inspection conducted to review an April 6, 1994 incident in which a patient received a dose of 10.4 Gray instead of the intended 6 Gray per fraction during a treatment with a high dose rate remote afterloader. The inspection also included a review of the quality management program pertaining to brachytherapy. <u>Results</u>: Of the areas inspected, no violations of NRC requirements were identified. However, one unresolved issue was identified regarding the definition of a misadministration as it relates to the brachytherapy incident (Section 3).

9405040085 940426 PDR ADOCK 03001391 C PDR

DETAILS

1. Persons Contacted

+James Jones - Associate Director *Julie Stark - Assistant Director *Philip Dobrin, M.D. - Assistant Chief of Staff *Lawrence Case - Radiation Safety Officer *Bernard Kelly - Assistant Radiation Safety Officer Glenn Glasgow, Ph.D. - Radiation Physicist Leonid Leybovich, Ph.D. - Radiation Physicist

+Present at entrance meeting held April 13, 1994. *Present at exit meeting held April 13, 1994.

2. Licensed Program

V.A. Edward Hines Jr., Medical Center (V.A. Hines or Licensee) is authorized by License No. 12-01087-07, a medical broad scope license, to use, in part, sealed sources of iridium-192 (Ir-192) in an Isotopen-Technik Dr. Sauerwein GmbH Model GammaMed III high dose rate remote afterloading device (GammaMed HDR) for interstitial and intercavitary treatment of cancer.

3. Brachytherapy Incident

The following is a summary from the interviews held with the Licensee's radiation physicists.

On April 6, 1994, a patient being treated for cancer of the larynx was scheduled to receive the first of two radiation treatments using a GammaMed HDR. Each treatment was to deliver a dose of 6 Gray (Gy), for a total dose of 12 Gy. Some time before the treatment, the radiation therapist correctly entered the treatment parameters into the GammaMed computer. The entry of these parameters was checked by another therapist and a radiation physicist. This check was documented on the licensee's HDR treatment form. At the time of the treatment, the therapist recalled the treatment parameters from the computer, and, by mistake, entered the date in the incorrect format, i.e., 040694 (June 4, 1994) rather than 060494 (April 6, 1994). She then checked the dwell times for each position, which were correct, and started the treatment. During the post-treatment quality assurance check, it was discovered that the treatment time was too long by a factor of 1.74. Further checking revealed that the date had been entered incorrectly, as discussed above. The long treatment time was accounted for by the data entry error (see Appendix A). This resulted in the patient receiving a dose of 10.4 Gy instead of the intended 6 Gy for the first planned treatment. As of the date of the inspection, the radiation oncologist had not yet decided whether to give a second treatment of 1.6 Gy, or to modify the written directive and cease treatment. The licensee notified Region III of this incident on April 8, 1994.

A NRC medical consultant has been contacted to review this incident and provide an opinion on the biological effects expected for the patient.

The incident is still under review by the NRC to determine if a misadministration, as defined in 10 CFR 35.2, occurred (URI 94001/01).

Regarding corrective actions, shortly after the incident, the radiation physics staff held a meeting where the HDR operating procedures, especially the data entry requirements, were reviewed. New, large signs were posted near the HDR console to remind the therapists of these requirements. Also, the HDR treatment form was modified to: (1) require the therapist to write the date in the same format as it must be entered into the computer, and (2) immediately before treatment, require another check of the treatment parameters (this is in addition to the first data entry check which is performed when all the treatment parameters are initially entered into the computer). Moreover, the licensee will begin using a manual, backup timer to check the agreement with the HDR timer.

One unresolved item was identified.

4. Quality Management Program (QMP)

The inspector performed a review of the licensee's OMP (HDR section only). Since the last inspection, nine HDR treatments have been performed. For each treatment, the written directive was prepared as required. A patient's identity was verified by asking the patient his/her name, and by comparing the individual with a photograph of the patient which is kept in the patient's chart. The treatment plans were formulated to deliver the prescribed dose as listed on the written directive, and were always checked by a second individual prior to treatment. The treatment planning was performed in accordance with the manufacturer's instructions and planning atlas. The entry of the treatment parameters into the computer was checked by a second individual prior to treatment. After the treatment, a check is performed to assure that the treatment was delivered as planned, to within $\pm 10\%$. The incident described above was the only example of an unintended deviation from the written directive. As discussed already, this deviation was immediately identified, and the licensee took prompt and effective actions to prevent recurrence. The incident and the corrective actions were thoroughly documented (see Appendix B).

No violations of NRC requirements were identified.

5. Exit Meeting

An exit meeting was held at the conclusion of the inspection with those individuals identified in Section 1 of this report. A summary of the inspection was discussed. The licensee did not identify as proprietary any of the materials provided to or reviewed by the inspectors.

APPENDIX A

Iridium-192 (Ir-192) has a half-life $(r_{\%})$ of 73.83 days. On February 2, 1994, the activity (A) of the Ir-192 source was 9.9 Curies (Ci). The activity of a radioactive sample at time τ is calculated from

$$A(\tau) = A(0) e^{-\frac{\ln 2}{\tau_{W}}(\tau)}$$
.

where A(0) is the initial activity. Thus, on April 6, 1994, 63 days later, the activity was 5.48 Ci, since

$$A(63) = 9.9 Ci e^{-\frac{1n2}{73.83}(63)} = 5.48 Ci.$$

On June 4, 1994, 122 days later, the activity would be 3.15 Ci, since

$$A(122) = 9.9 Ci e^{-\frac{\ln 2}{73.83}(122)} = 3.15 Ci.$$

The treatment time would be increased by a factor of 1.74, since

$$\frac{5.48}{3.15} = 1.74$$
.

APPENDIX B

04/11/94 15:07

RIII 7085151078

00/001/001A CC 00/AS - Larry Case WHA Date: April 11, 1994 001

TIFICATION OF EVENT OR UNUSUAL OCCURRENCE PNO-III-94-22

Or Biminary notification constitutes EARLY notice of events of POSSIBLE safety or public interest significance. The information is as initially received without verification or evaluation, and is basically all that is known by the Region III staff on this date.

-1--

Facility Veterans Affairs Medical Center Hines, IL 60141

Licensee Emergency Classification General Emergency site Area Emergency Alert Unusual Event X Not Applicable

License No. 12-01087-07 Docket No. 030-01391

Subject: RADIATION THERAPY ERROR

On April 8, 1994, the licensee informed Region III (Chicago) that an error had occurred during the radiation treatment of a patient using a Gamma Med II HDR unit and an iridium-192 sealed source. The patient was scheduled to receive a series of two treatments of 600 rads (6 gray) for a total dose of 1,200 rads (12 gray). Because of an error in the treatment parameters, the patient received 1,000 rads (10 gray) in the first treatment on April 7, 1994.

On April 6, 1994, during treatment planning, the licensee entered the wrong date into the HDR unit. The date, "4-6-94," was incorrectly entered as "6-4-94." The treatment time is based on the computed strength of the iridium-192 source. Since the iridium-192 source with a 74-day halflife would have a lower strength on 6-4-94, the HDR unit increased the calculated treatment time resulting in the greater than intended dose.

The licensee intends to modify the written directive for the treatment to compensate for the error. For the second treatment, the dose will be 200 rads (2 gray). Therefore, the patient will receive the intended total dose.

Region III will conduct a special inspection this week to evaluate the circumstances surrounding the error. An NRC medical consultant has also been contacted regarding the case.

The State of Illinois has been informed. (Illinois is an Agreement State but federal facilities remain under NRC jurisdiction.) The information in this preliminary notification has been reviewed with licensee management.

Region III was notified of this event on April 8, 1994. This information is current as of 1:00 p.m. on April 11, 1994.

CONTACT:

Thomas Young 708/829-9835

Roy Caniano 708/829-9804





REPORT

of an error that occurred during irradiation of a patient with the Gamma-Med IIi HDR remote afterloading unit.

According to the treatment plan, the radiation therapist Ms. Reynolds keyed into the Gamma-Med computer the treatment parameters, which were verified by another radiation therapist, Ms. Johnson, and by the radiation physicist, Dr. Leybovich, and stored in the Gamma-Med computer.

However, when Ms. Johnson recalled the treatment parameters from the computer memory, she inadvertently keyed in a wrong date, 4.06.94* (June 4, 1994) instead of the correct date, 6.04.94* (April 6, 1994). She checked the dwell times for each position and the total time which were correct, and started the treatment.

Because the Gamma-Med computer adjusts the FAK and AKT parameters automaticly, these factors have been changed by the computer according to the wrong date. These parameters are used by the treatment computer to modify the planned (keyed) treatment time according to the source decay. Thus, the treatment computer used a greater modifying factor (3.17) instead of the required modifying factor (1.83).

This resulted in administration to the patient a dose of 10.39 Gy instead of planned 6.0 Gy per fraction (6.0 Gy x (3.17/1.83)= 10.39 Gy). However, the total prescribed dose (12.0 Gy in two fractions) was not exceeded because the error occurred during administration of the first fraction.

Louid Leybaring

Leonid B. Leybovich, Fn.D. Assistant Professor

NB: Filed W.D. Karry lare - RSO - \$ S. Lulla, MID - Que Convertes 4-8-94

*The Gamma-Med treatment computer uses a European notation.

Note: I believe this constitutes a locordable event as the last frection has not been Juer. D. Reday will see the patient again in two meeks, exomination to tumor repressing and decide then about giving the 1667. Herm P. Hagar, A.D. 4-8-94

	Loyola University	Edward Hines V.A. Hospital	L/H:	YEAR:	NUMBER:				
	Medical Center 2160 South First Ave.	Mail Route #114-B Dept. of Radiotherapy	H	94					
	Maywood, IL 60153	Hines, IL 60141							
2	(708) 531-3930	(708) 343-7200	SOCIAL SECURI	TY NUMBER					
- D	PATIENT'S NAME - LAST,	REDAY Pauling RADIOTHERAPIST							
	DATE DIAG	NOSIS	AGE	SEX	RACE	PLACE			
	Apr. 6, 1994 2	Suprachattic laryn	× Z		1000	TREAT-			
	TREATMENT PLANNING		BILLING CO	ODES:		-MENT TAPE			
	NO1.		777	77781 - 1 to 4 source positions/catheters					
	PRESCRIPTION BY: DOGULINO			77782 - 5 to 8 source positions/catheters					
	DATE: 4/6/	74	777	784 - over 12 sourc	e positions/catheters				
	PRESCRIDED DOSE (Gy)	12 6 at	10 in tissue (mm)	DOSE/FRAC	TION 6	(Gy)			
	NUMBER OF FRACTIONS	2NUMBER OF WEEKS	2 п	REATMENT NUM	BEROF2				
	INITIAL DATE	TREATMENT	ELAPSED		CURRENT ACTIVITY	(
	Feb. 2. 1994	April 6, 1994	63		. 5536 × 9.9 Q	= 5.48 6			
	INITIAL ACTIVITY	1.9 CL FAK TODAY	183	% AKT	TODAY 54	\$			
				in the second					
	PATIENT NUMBER	APPLICATOR <u>2 sophageal</u>	NO. OF CYCI	ES	PROGRAM STORED	1			
-	CHANNEL NO.	OF SKIP-STEP	80	mm IRRAD	IATED LENGTH 70				
	TIME SET (seconds):	. 56	Position 11_						
	Politic								
113	112 POBLE	~ 28							
S	A Positio	25							
ET APR	8 1994 E Positio	m 04							
E nr	POSITION OF POSITIO	2 6				1.51			
10 KI	CUCIVLI A Position	30 06 <u>2 0</u>							
B	A Positie	0n 07 <u>21</u>				1. 1. 1.			
The	TOTTS Positi	06.08_26	Position 19						
	Positi	00 09760 8le	Position 20						
	Positi	oe 100		TOTAL CHANNEL	The 26 Osee	-			
	TOTAL TIME-	SUM OF ALL CHANNELS	-		Al Al				
	PLANNED BY. NO.	rid duberich	CONFIRMED BEI	FORE TREATMEN	THY W Range day	Jan 2015			
	DATA ENTRY AND TRE	DATA ENTRY AND TREATMENT BY RUN ALL TO DATA CONFIRMED RECORE TREATMENT BY AND TREATMENT BY							
	TOTAL TIME SET 260 × FAK TODAY 1.83 × REAL USE TIME DELIVERED 47578								
	FROM PRINT OUT, TOTAL TIME SET 260 × FAK TODAY 3.17 = REAL USE TIME DELIVERED 82430								
0	EXPOSURE RATE ON THE PATIENT SURFACE AFTER TREATMENT = 0.01 MR/HR								
3	SURVEY METER USFO: Ludlum Model 3 SN: 25910 SURVEYED BY: Lever's deposit								
5	Reyed in instead of 6.04.94 (European notation), the patient receiversed:								
	SKONATURE	long deglor	06.	DATE 41	477 124	I LAULAR.			
N	10.04 64 148	run of pranned of	e ay						

To: All Staff With Gamma Med Responsibilities



Fm: Glenn P. Glasgow, Ph.D.

Dt: 4-12-94

Re: Revision in QA Procedures For Gamma Med Patients

Two items require revision in our QA procedures for the Gamma Med.

(1) We had an incident last week in which, in recalling a stored program, the date was inadvertently entered incorrectly. April 6, 1994 (6.04.94) was entered as June 4, 1994 (4.06.94). While we performed all established CA checks as the treatment parameters were entered into the computer and stored, when the program was recalled the displayed dwell times were checked, but the operator failed to check a second time the displayed FACTOR and ACTIVITY values, which were now displayed for June 4 rather than for April 6.

We need to perform a through check of all displayed parameters for each treatment immediately before pressing the START button. I am designing a new form that will allow explicit check-off of each parameter not only at the time of entry but also at the time of recall for stored programs.

(2) We have also discovered, during some QA testing, that it is possible to make an error, probably without realizing it, in entering the data for pre-stepping. In the blank marked SKIP-STEP, recall that to set a pre-step distance of 40 mm with a step distance of 2 mm, the proper entry is DIS: 4002. However, if you enter this as DIS: 402, omitting a zero, the unit will interpret this as a 4 mm skip with a 2 mm step. Hence, it is very important to enter this correctly. The computer will not accept leading zeros. Hence, if you entered 0402 for a 4 mm skip with a 2 mm step, the computer will not accept these values.

Likely clinical values are skips of integer millimeters, e.g., 10 mm, 20 mm, with steps of either 5 mm or 10 mm steps. These are properly entered as Dis: 1005, 1010, 2005, 2010.

If there are no skips, the proper entries are 5, 10, 15, 20 for steps of 5 mm, 10 mm, 15 mm, and 20 mm respectively.

(3) Regarding actions to take to help us recognize if we have some error that we fail to recognize, I am obtaining an external timer with an alarm. We will be developing some procedures to use this timer as an additional safety check mechanism.

(4) I will review these new procedures with you at a short conference that I will schedule later this week.

APR. 14 '94 16:03 RADIOTHERAPY VA-LUMC

GPG 4-13-94

INSTRUCTIONS FOR ENTERING A FROGRAM INTO MEMORY AND RECALLING A PROGRAM GENERAL POINTS:

*Only 23 programs may be stored in memory. *Each program can contain up to 12 pages or channels. *Dwell times must not exceed four digits. *Position distances and number of cycles are limited to two digits. *Total irradiation time is limited to six digits. *Skip distances are limited to three digits

ENTERING A PROGRAM INTO MEMORY:

- Enter the desired program following the general instructions for keying in a program. Check the accuracy of the program before entering into memory.
- 2. At the end of keying in data, you will be on page "O". Assign a desired number to the program, e.g., 1, 2, 3 ... 11, etc. However, under PROGRAM NO., enter this program number as a DOUBLE ENTRY with a "O" separating the entries, if it is a single digit number. For example, Program 1 will be entered as Program No. 101, Program 11 as Program No. 1111.
- 3. Press PROGRAM AUTOMATIC. The program is now stored in memory.

4. Record information about program in the "Program Log Book".

RECALLING A STORED PROGRAM:

- Enter the heading data specific for your treatment date. For example, Date [REMEMBER Day-Month-Year], Patient No.,
- 2. With the console "ON" and in the "READY" mode, i.e., ready to have data entered, enter the desired program number on page 0 under Program No., e.g., 1, 2, 3 ... 11. (NOTE: Do not use double entry when retrieving programs. Identify from the Log Book, the stored program number you desire to execute.
- 3. Press PROGRAM AUTOMATIC. "FAK" will clear to 0. The total time will be displayed. Verify the Total Time setting per log book. Press Time Factor Auto. The ACT and FAK must be for the current date. Press Page No., number 1 should appear next to "Page" on the monitor. Have a second operator, physicist, dosimetrist, or physician cross-check the data recalled from storage. Go through all remaining pages and have a second person cross-check the data (if any). When return to 0 page, have a second person cross-check ACT and FAK.
- 4. Press START and SOURCE IN IRRADIATION POS simultaneously to execute treatment. As program runs, observe data display screen to verify that proper page number, stepping distance, position numbers, and dwell times are being executed.

P.5

REMEMBER — REMEMBER ENTER THE DATES IN THE EUROPEAN NOTATION: DAY - MONTH - YEAR For example: April 6, 1994 is entered as

060494

4-13-94

PRESCRIPTION-TREATMENT FORM CHECKLIST PROCEDURES

G4PL 4-13-9

Please note the changes made in our Prescription/Treatment Form.

(1) Please note that t: eatment date, near the center of the form, has been labelled o clearly denote the european formal of month-day-year, for entry into the computer. It is imperative that this format be used to enter the date into the computer. Boxes have been added for entering these dates.

(2) Please note that the skip-step data entry has also been clearly labelled and that boxes have been added for entering these numbers. For example, if NO skip distance is to be used, and only a step, enter the step distance, in millimeters. The computer will not accept leading zeros. Hence, 5 mm will be entered in the last box as 5; 10 mm will be entered in the last two boxes as 1 0.

If a skip distance is planned, e.g., 120 mm, with a 10 mm step, enter in the five boxes as 1 2 0 1 0. If a 20 mm skip is planned, with a 10 mm step, do not use the first box and enter data in the last four boxes as 2 0 1 0. Do not us a leading 0 in the first box. If the plan is a 20 mm skip with 5 mm steps, this will be entered as 2 0 0 5 in the last four boxes.

(3) Check boxes have been added to the form for the use of those actually checking work. After the dwell times have been entered by the dosimetrist/physicists (following a second check of this work by a second individual), radiation therapist #1 will key the parameters into the computer under the supervision of radiation therapist #2. Therapist #2 will physically check the appropriate first row of boxes. Now, there are often delays in the actual treatment. If the entered parameters are NOT stored, I still want a check, using the second row of boxes, by a member of the physics/dosimetry staff IMMEDIATELY before treatment. Note there is a new signature line for this check. If the data is STORED, perform this third check AFTER the data is recalled, but IMMEDIATELY before treatment. Be certain to physically check each parameter to make sure that no errors have been made during the recall of the stored program.

(4) Near the bottom of the page, use the date line after the "Surveyed By" line to date the page immediately after treatment. Use the last date line to indicate the date that any additional work is done after treatment to validate the conclusion that the treatment was delivered correctly to within 10%.

GAMMA ME Loyola Universit Medical Center 2160 South First	D HDR BRACHYTHERAPY PLA y Edward Hines V.A. Hospital Mail Route #114-B Dest of Bedictherapy	NNING & TR	EATMENT R	ECORD 412.94 NUMBER:		
Maywood, IL ((708) 531-3930	60153 Hines, IL 60141 (708) 343-7200	SOCIAL SECURITY NUMBER				
PATIENT'S NAME	- LAST, FIRST, MIDDLE					
DATE	DIAGNOSIS	AGE	SEX	RACE		
TREATMENT PLAN	I NNING:	BILLING CODES: 				
PRESCRIPTION BY						
DATE	1996-1996 (1997) - 1995 (1996) - 1996 (1996) - 1996 (1997) - 1996 (1996) - 1996 (1996) - 1996 (1996) - 1996 (19 1996 - 1996 (1997) - 1996 (1996) - 1996 (1996) - 1996 (1996) - 1996 (1996) - 1996 (1996) - 1996 (1996) - 1996 (1					
PRESCRIBED DOSE	£ (Gy)	(mm)	DOSE/FRACTION	(Gy)		
NUMBER OF FRAC	TIONSNUMBER OF WEEKS	TRE	ATMENT NUMBER	OF		
NITLAL DATE	TREATMENT DATE MORTH ; YEAR		ELAPSED DAYS	CURRENT BY CHART		
			_			
	homenand lanaread lances lances have					
NITLAL ACTIVITY	FAK TODAY	% 🗆 🗆	AKT TODAY_	% 🗆 🗆		
PATIENT NUMBER	APPLICATOR NO). OF CYCLES		CORAM STORED		
				CORAM STORED		
CHANNEL NO	OF 🗆 🗌 📃 📃		mm GRRADIATED	LENGTH		
TIME SET (Laconda)	SKD2	STEP	and a second second second second			
	Position 01	Position 11	[] []			
	Position 02	Position 12				
	Position 03	Position 13				
	Position 04	Position 14				
	Position 05	Position 15				
	Position 06	Position 16				
	Position 07	Position 17				
	Position 08	Position 18				
	Position 09	Position 19				
	Position 10	Position 20				
TOTAL TIME-SUM	OF ALL CHANNELS	TOTAL	CHANNEL TIME SE	τ□□		
LANNED BY		CONFIRMED BEFOR	E TREATMENT BY			
DATA ENTRY BY:_	DATA	ENTRY CONFIRME	EL BY:			
ATA PARAMETER	CONFIRMED IMMEDIATELY BEFORE TREAT	MENT BY:	an a	an analysis and a strain all an an an and a strain and a strain and		
OTAL TIME SET_	× FAK TODAY	= REAL USE TIME DELIVERED				
ROM PRINT OUT,	TOTAL TIME SET × FAK TOI	YAC	= REAL USE TEM	E DELIVERED		
HOVEN METER (IN THE PATIENT SURFACE AFTER TREATMEN	/[e				
OWTHIN 104 T	WIS TO FATMENT WAS DELIVERED	SU/	RVEYED BY	DATE:		
Con arrite	THE TREATMENT WAS DELIVERED, 43 PLANY	VED.		REVISED.		
DIGNATURE	an a		DATE	412.94CKG		

GPG 4-13-94

P.2.

GAMMA MED BACK-UP TIMER CALCULATIONS

INITIAL DATE OF ACTIVITY	TREATMENT DAY	DATE MONTH	YEAR	ELAPSED DAYS	CURRENT ACTIVITY BY CHART	
INITIAL ACTIVITY		FAK TODAY		% AKT 1	ODAY	8
PLANNED TIME POSITION 1		FAK × Today		REAL = POSIT	TIME TON 1	
Calculated	By:			Date:		

Note: The back-up timer should be set by someone other than radiation therapist. The Physicists or dosimetrists performing the parameter check immediately before treatment should set the clock. However, the radiation therapist operator will start the clock.