

DIRECTORS

R. V. BRIDWELL
U. GORDON COLSON
ROBERT L. GIBSON
W. HOEFFDING, M. D.
MRS. MARY ELLEN HONNOLD
J. M. INGALLS, M. D.
EDWARD H. JENISON
C. A. MCCLELLAND, M. D.
WILLIAM MILBOURN
MRS. R. S. PEARMAN, JR.
J. R. SHACKELFORD, M. D.
G. H. SPRAGUE, M. D.
R. G. WIESE
C. E. WILIS

HOSPITAL AND MEDICAL FOUNDATION OF PARIS, INC.

A NON - PROFIT ORGANIZATION

PARIS, ILLINOIS 61944

TELEPHONE 217-465-4141

W. W. VONDERLAGE
ADMINISTRATOR

Aug. 7, 1981

United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL. 60137

D.G. Wiedman, Acting Chief
Materials Radiation Protection, Section I.

Re: N.R.C. By-Product Material License No. 12-14142-01

Dear Sir,

The following statement was affirmed under oath, concerning the routine safety inspection conducted by Mr. W.P. Reichhold at our institution on June 22, 1981. I wish to affirm the following:

Concerning Item 1. and Item 2., of Appendix A, dealing with an un-authorized user and non-designated Safety Officer. This is a result of the death of Dr. Hogan and his replacement by the undersigned, which was a matter of some urgency. The delay in applying for the amendment has been caused in part by an oversight, and in part by the resignations or deaths of key personnel at both institutions at which I completed my residency. The necessary preceptor statements as well as other documentation required has been submitted to your department. Compliance on these two Items will be achieved as soon as approval has been granted.

Item 3. refers to unsatisfactory constancy checks for the dose calibrator, which we were performing on a weekly basis. We have since changed our procedure and now perform that test on every day in which the dose calibrator is used. This will be reviewed on a quarterly basis at our meetings of the Radiation Safety Committee, to insure compliance.

Item 3. also refers to annual accuracy tests. We are submitting copies of these tests dating back to 1977, and hope that these will meet with your requirements.

Item 4. deals with Radiation Safety Committee meetings, in which we have been slow to organize. I accept responsibility for this and will insure that this will not occur again.

8109170310 810902
NMS LIC30
12-14142-01 PDR

AUG 13 1981