



**SHENANGO VALLEY
OSTEOPATHIC HOSPITAL**

2200 MEMORIAL DRIVE EXTENDED

FARRELL, PENNSYLVANIA 16121

Docket No 30-9912

September 2, 1981

John D. Kinneman, Chief Materials
Radiological Protection Section
Technical Inspection Branch
U.S. Nuclear Regulatory Commission
631 Park Avenue
King of Prussia, PA 19406

RE: Docket No. 30-09912
License No. 37-15857-01

Dear Sir:

In reference to your letter dated August 28, 1981 regarding the routine safety inspection conducted by Mr. J. Nicolosi on July 7, 1981, please note that syringe shields are now being used for the preparation of patient doses and will continue to be used in the future. Two new shields have recently been purchased which allows 360° visibility for easier use.

Sincerely,

C. E. Clary
Hospital Administrator
Shenango Valley Osteopathic Hospital
2200 Memorial Drive Extension
Farrell, PA 16121

CEC:mvk

8110220198 811008
NMS LIC30
37-15857-01 PDR

AREA CODE 412-981-3500