

Appendix A

NOTICE OF VIOLATION

Dearborn County Hospital

License No. 13-17327-01

As a result of the inspection conducted on October 22, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. License Condition No. 14 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in application dated May 6, 1977.

Page 2 of the above referenced application states the Radionuclide Radiation Safety Committee shall meet at least four times per year.

Contrary to the above requirement, it was determined through statements by licensee representatives and the NRC inspectors review of records that this condition is not being met. Specifically, your Radiation Safety Committee met on October 10, 1979, January 15, 1980, and February 12, 1981.

This is a Severity Level V violation (Supplement VII).

2. License Condition No. 14 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in application dated May 6, 1977.

Item No. 10 of the above referenced application states, a Texas Nuclear Log Series Survey Meter Model 9121 with a sensitivity range of .2 to 2000 mR/hr will be ordered and used to conduct routine area surveys.

Contrary to the above requirement, it was determined through statements by licensee representatives and the NRC inspectors review of records that this condition is not being met. Specifically, your Minimonitor II, Serial No. 8115 has a sensitivity range of 0-10 mR/hr.

This is a Severity Level V violation (Supplement VII).


Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written

8111200694 811112  
NMS LIC30  
13-17327-01 PDR

statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

Dated

11/2/81

  
\_\_\_\_\_  
D. J. Sreniawski, Chief  
Materials Radiation Protection  
Section 2