

A Regional Medical Center

2525 Glenn W. Hendren Drive • Liberty, Missouri 64068 • (816) 781-7200

Lie. File

Mr. R. J. Caniano, Chief Nuclear Materials Safety Section 2 United States Nuclear Regulatory Commission Region III 799 Roosevelt Road Glen Ellyn, Illinois 60137

RE: Liberty Hospital License No. 24-16178-01

March 5, 1990

Dear Sir:

Upon receipt of our Notice of Violation, we are hereby submitting this letter of response as required by 10 CFR 2.201. The following items are the responses to our violations:

- 1.) Special attention will be given to insuring that our Medical Isotope Committee will meet once each quarter. The next meeting is slated for April 5, 1990, and it will be made a practice to set the date for the following meeting at each session. Our physicist makes quarterly reports, and this will be coordinated with our Medical Isotope Committee meeting. Administration and the current Radiation Safety Officer are both acutely aware of the need for continuity in our Safety meetings.
- 2.) Subsequent to the inspection, the enclosed records were produced, indicating that the dose was assayed as required by License Condition No.22 stating that each patient dose must be assayed in the dose calibrator prior to administration.

Please inform us if there is any other information that we need to submit, and we will promptly comply.

Sincerely:

Vicki L. Cooper

Director of Radiology Services

Vicki L. Cooper

VLC/ph

