

Health Physics Sample Records

Sample ID: 142473

Type: RLSSS

Type Detail: SAND

Status: APPROVED

Rasterized:

Description: 140603BD 2, CULVERT #3

Common

Resp Org: P4000 - PROJECT MANAGEMEN

Cross Ref Sample ID:

Priority: N

Contact: COOPER, TERRY L

Survey No:

Reference ID:

Accounting Subcode:

Unknown Mat: NO

Hazardous Mat: NO

Comments: COUNT TO MDA

Reviewed: DYNES, WILLIAM J

Date: 06/05/14 07:01

Comments:

Origin

Mat. Size/Volume: 1642 GRAMS

Docmnt/MO/Act #:

Proposed Ship Date: 00/00/00 00:00

Description: CULVERT 3

Collection

Sample Tech: DYNES, WILLIAM J

Sample Date: 06/03/14 07:57

Count Time: 900 (secs.)

Tritium Req: No

Sample Size/Volume: 1642 GRAMS

Instructions: COUNT TO MDA

Comments: COUNT TO MDA

RLSSS

Exempt: No

Evaluated: No

Chem Notification: No

State Permit: No

Offsite Analysis: No

Effluent Permit: No

Container Surveyed: No

Cleared For Release: Yes

Container Evaluated: Yes

Form 212 Ready:

Location

Unit: NON-UNIT RE

Area: MS

Elev: 100

Room:

Depth: (in.)

Equip ID:

Stored By: ,

Date: 00/00/00 00:00

Sample Loc:

Other:

Material Loc:

Health Physics Sample Records

Sample ID: **142473**

Type: RLSSS

Type Detail: SAND

Status: APPROVED

Rasterized:

Description: 140603BD 2, CULVERT #3

Analysis

Geometry	MCA ID	Quant.	Units	Method	FP/AP	Lab	Analysis By	Date	Mod ID	Mod Date
C_1L_marinelli_sand_	30377	1642	GR		.000		PARK, MICHAEL C	06/05/14 16:28		
Isotope	Mode	Activity	Activity Uncert.	LLD/MDA	% Abund.	Atom. No.	Comments	Mod Id	Mod Date	
0 K-40	NT	0.00001757	1.66E-6	<input type="checkbox"/>	92.442%	40				
0 PB-212	NT	4.7E-7	1.173E-7	<input type="checkbox"/>	2.473%	212				
0 BI-214	NT	3.149E-7	7.37E-8	<input type="checkbox"/>	1.657%	214				
0 PB-214	NT	6.517E-7	1.796E-7	<input type="checkbox"/>	3.429%	214				
0 BE-7	AP	8.3663E-7		<input checked="" type="checkbox"/>	0.000%	7				
0 MN-54	AP	1.1637E-7		<input checked="" type="checkbox"/>	0.000%	54				
0 MN-56	AP	8.495E-8		<input checked="" type="checkbox"/>	0.000%	56				
0 CO-57	AP	7.9134E-8		<input checked="" type="checkbox"/>	0.000%	57				
0 CO-58	AP	8.0125E-8		<input checked="" type="checkbox"/>	0.000%	58				
0 CO-60	AP	4.3554E-8		<input checked="" type="checkbox"/>	0.000%	60				
0 SB-124	AP	8.2179E-8		<input checked="" type="checkbox"/>	0.000%	124				
0 SB-125	FP	2.3222E-7		<input checked="" type="checkbox"/>	0.000%	125				
0 CS-134	FP	9.3676E-8		<input checked="" type="checkbox"/>	0.000%	134				
0 CS-137	FP	9.649E-8		<input checked="" type="checkbox"/>	0.000%	137				

Analysis Review

Re-Analyze: No

Lic Rad Mat Present: No

Reviewed: COOPER, TERRY L

Date: 06/10/14 14:26

Hp Eng Rvw Req'd: Yes

Rad Mat Present: Yes

HP Notified Name:

Comments:

Health Physics Sample Records

Sample ID: 142473

Type: RLSSS

Type Detail: SAND

Status: APPROVED

Rasterized:

Description: 140603BD 2, CULVERT #3

Review/Approve

HP Eng Review

Reviewed: COOPER, TERRY L

Date: 06/10/14 14:26

Comments:

Disposition

Disposition: OTHER

Result: Sample contains only natural isotopes

Remarks:

Final Approval

Approved: COOPER, TERRY L

Date: 06/10/14 14:26

Comments:

```

*****
*          SONGS                      HEALTH PHYSICS REPORT          5-JUN-2014 16:25  *
*-----*
*          >>>> SAMPLE PARAMETERS <<<<<          *
*
* Title      : 140603BD 2, CULVERT #3          *
*
* ID         : 30377                          Acquisition date: 5-JUN-2014 16:22:28. *
* Start Time: 5-JUN-2014 16:22:00.           Counting System   : U2GSP          *
* Stop Time  : 5-JUN-2014 16:22:00.           IHPS Sample Number: 142473         *
* Quantity   : 1.64200E+03 GR                  Analysis Library  : HP_NATURAL      *
* Geometry   : C_1L_marinelli_sand_1          Operator: MICHAEL PARK             *
*-----*
*          >>>> C #41-TP11750A PARAMETERS <<<<<          *
*
* Elapsed Live Time : 0 00:03:00.60           Calibration       : 5-JUN-2014 08:14: *
* Elapsed Real Time : 0 00:03:00.64           Energy Offset    : -1.39807E-01       *
* Deadtime          : 0.0%                    Energy Slope     : 4.99820E-01         *
* Activity Mult Factor: 1.00000E+00           Energy Tolerance: 1.00000          *
* Half Life Ratio   : 8.00000                 Sensitivity      : 5.00000             *
*
*****

```

Post-NID Peak Search Report

It	Energy	Area	Bkgnd	FWHM	Channel	Left	Pw	%Err	Fit	Nuclides
1	238.82	46	14	1.89	478.07	472	15	25.0	3.70E+00	PB-212
1	295.56	24	3	1.85	591.59	587	13	27.6	1.00E+00	PB-214
1	351.84	23	8	1.50	704.17	699	12	33.5	1.06E+00	PB-214
1	609.20	17	0	1.88	1218.99	1215	8	23.4	4.43E-01	BI-214
1	1460.86	114	0	2.07	2922.26	2916	12	9.5	9.36E-01	K-40

Summary of Nuclide Activity
Sample ID : 30377

Page : 2
Acquisition date : 5-JUN-2014 16:22:28

Total number of lines in spectrum 5
Number of unidentified lines 0
Number of lines tentatively identified by NID 5 100.00%

Nuclide Type : natural

Nuclide	Hlife	Decay	Uncorrected uCi/GR	Decay Corr uCi/GR	Decay Corr 1-Sigma Error	1-Sigma %Error	Flags
K-40	1.28E+09Y	1.00	1.757E-05	1.757E-05	0.166E-05	9.46	A
PB-212	10.64H	1.00	4.690E-07	4.700E-07	1.173E-07	24.95	A
BI-214	19.90M	1.07	2.940E-07	3.149E-07	0.737E-07	23.42	A
PB-214	26.80M	1.05	6.192E-07	6.517E-07	1.796E-07	27.56	A
			-----	-----			
		Total Activity :	1.895E-05	1.901E-05			

Grand Total Activity : 1.895E-05 1.901E-05

Flags: "K" = Keyline not found
"E" = Manually edited

"M" = Manually accepted
"A" = Nuclide specific abn. limit

Unidentified Energy Lines
Sample ID : 30377

Page : 3
Acquisition date : 5-JUN-2014 16:22:28

None

Flags: "T" = Tentatively associated

Interference Report
Sample ID : 30377

Page : 4
Acquisition date : 5-JUN-2014 16:22:28

No interference correction performed

Nuclide Type: natural

Nuclide	Energy	Area	%Abn	%Eff	Uncorrected Decay Corr		1-Sigma %Error	
					uCi/GR	uCi/GR		
K-40	1460.83	114	10.67*	5.560E-01	1.757E-05	1.757E-05	9.46	
PB-212	74.81	-----	10.50	1.570E+00	-----	Line Not Found	-----	
	77.11	-----	17.70	1.661E+00	-----	Line Not Found	-----	
	87.19	-----	6.27	1.983E+00	-----	Line Not Found	-----	
	90.13	-----	1.86	2.055E+00	-----	Line Not Found	-----	
	115.12	-----	0.59	2.376E+00	-----	Line Not Found	-----	
	238.58	46	43.60*	2.054E+00	4.690E-07	4.700E-07	24.95	
	300.03	-----	3.34	1.843E+00	-----	Line Not Found	-----	
	BI-214	76.86	-----	0.58	1.651E+00	-----	Line Not Found	-----
		79.29	-----	0.97	1.741E+00	-----	Line Not Found	-----
		609.31	17	46.10*	1.112E+00	2.940E-07	3.149E-07	23.42
665.44		-----	1.56	1.023E+00	-----	Line Not Found	-----	
768.35		-----	4.88	8.873E-01	-----	Line Not Found	-----	
806.16		-----	1.23	8.454E-01	-----	Line Not Found	-----	
934.04		-----	3.16	7.310E-01	-----	Line Not Found	-----	
1120.27		-----	15.00	6.242E-01	-----	Line Not Found	-----	
1155.18		-----	1.69	6.105E-01	-----	Line Not Found	-----	
1238.11		-----	5.92	5.846E-01	-----	Line Not Found	-----	
1280.95	-----	1.47	5.747E-01	-----	Line Not Found	-----		
1377.66	-----	4.02	5.602E-01	-----	Line Not Found	-----		
1385.30	-----	0.78	5.595E-01	-----	Line Not Found	-----		
1401.48	-----	1.39	5.583E-01	-----	Line Not Found	-----		
1407.97	-----	2.48	5.578E-01	-----	Line Not Found	-----		
1509.22	-----	2.19	5.568E-01	-----	Line Not Found	-----		
1583.22	-----	0.72	5.628E-01	-----	Line Not Found	-----		
1661.26	-----	1.15	5.753E-01	-----	Line Not Found	-----		
1729.58	-----	3.05	5.918E-01	-----	Line Not Found	-----		
1764.49	-----	15.90	6.023E-01	-----	Line Not Found	-----		
1847.41	-----	2.12	6.331E-01	-----	Line Not Found	-----		
PB-214	53.17	-----	1.10	5.203E-01	-----	Line Not Found	-----	
	74.81	-----	6.52	1.570E+00	-----	Line Not Found	-----	
	77.11	-----	11.00	1.661E+00	-----	Line Not Found	-----	
	87.19	-----	3.88	1.983E+00	-----	Line Not Found	-----	
	90.13	-----	1.15	2.055E+00	-----	Line Not Found	-----	
	241.92	-----	7.46	2.043E+00	-----	Line Not Found	-----	
	258.94	-----	0.55	1.981E+00	-----	Line Not Found	-----	
	295.09	24	19.20*	1.857E+00	6.192E-07	6.517E-07	27.56	
	351.87	23	37.10	1.686E+00	3.399E-07	3.577E-07	33.48	
	785.83	-----	1.09	8.674E-01	-----	Line Not Found	-----	
839.00	-----	0.59	8.122E-01	-----	Line Not Found	-----		

Flag: "*" = Keyline

Nuclide	Bckgnd Sum	Energy (keV)	MDA (uCi/GR)
BE-7	5.	477.61	8.3663E-07
MN-54	3.	834.83	1.1637E-07
MN-56	1.	846.81	8.4950E-08
CO-57	11.	122.06	7.9134E-08
CO-58	1.	810.79	8.0125E-08
CO-60	0.	1332.50	4.3554E-08
SB-124	2.	602.73	8.2179E-08
SB-125	3.	427.88	2.3222E-07
CS-134	3.	604.71	9.3676E-08
CS-137	2.	661.66	9.6490E-08