



Central Connecticut State University

A Campus of the Connecticut State University

1615 Stanley Street P.O. Box 4010 • New Britain, CT 06050 - 4010

January 4, 1993

Mr. Francis M. Costello, Chief
Research, Development &
Decommissioning Section
Division of Radiation Safety
and Safeguards
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road
King of Prussia, Pennsylvania 19406
Dear Mr. Costello:

Subject: Response to Letter January 14, 1993
Docket #030-12001, License #06-16975-01

In response to your letter of January 14, 1993, including Appendix A:

1. I have taken corrective steps for the cited violation: testing the sources for contamination. One is was on October 21, 1992 and the second text on January 19, 1993. These are recorded and show no leakage.
2. Future corrective steps will be taken such that the test for leakage will be done at least once every six months.
3. I shall be happy to send you the report whenever you request it for the leakage.

If you have any questions about this matter, please let me know.
Thank you very much for your consideration.

Yours truly,

Ali A. Antar, Ph.D.
Professor of Physics
Radiation Safety Officer

9303090163 930222
PDR ADDCK 03012001
C PDR



An Equal
Opportunity
Employer