

*CONTROL BLOCK:

--	--	--	--	--	--

 (1)

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

0	1	A	L	E	R	F	3	2	0	0	-	0	0	0	0	0	-	0	0	3	4	1	1	1	1	4			5
7	8	9	14					15	25										26	30					57	CAT	59		
LICENSEE CODE		LICENSE NUMBER										LICENSE TYPE																	

CON'T

REPORT SOURCE: 01 L 6 0 5 0 0 0 2 9 6 7 1 1 0 7 8 1 8 1 2 0 4 8 1 9

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

0 2 | During refueling outage, while performing SI 4.8.B.4-3, (reactor building vent
0 3 | monitoring system calibration) continuous air monitor 3-RM-90-250 failed to pass
0 4 | on gas and particulate channels. (T.S. 3.8.B-6, 3.8.B-8) There was no danger to
0 5 | the health or safety of the public. There were no previous similar events.
0 6 |
0 7 |
0 8 |

09		SYSTEM CODE		CAUSE CODE		CAUSE SUBCODE		COMPONENT CODE						COMP. SUBCODE		VALVE SUBCODE																													
7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46	47					
09		M C		E		G		I N S T R U						Y		Z		EVENT YEAR		SEQUENTIAL REPORT NO.		OCCURRENCE CODE		REPORT TYPE		REVISION NO.		ACTION TAKEN		FUTURE ACTION		EFFECT ON PLANT		SHUTDOWN METHOD		HOURS		ATTACHMENT SUBMITTED		NPRD-4 FORM SUB.		PRIME COMP. SUPPLIER		COMPONENT MANUFACTURER	
09		M C		E		G		I N S T R U						Y		Z		8 1		0 6 6		0 3		L		0		A		Z		Z		Z		0 0 0 0		Y		N		L		N 3 0 5	

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

1 0 | Component failure. Chemical Lab started sampling program. The NMC model CRA-14BF

1 1 | gas count rate card was replaced and the continuous air monitor was functionally

1 2 | tested and returned to service. This is considered a random failure and no further

1 3 | recurrence control is required.

1 4 |

FACILITY STATUS (1) H (2) 30
 % POWER (3) 0 (4) 0 (5) 0 (6) 29
 OTHER STATUS (7) NA
 METHOD OF DISCOVERY (8) B (9) 31
 DISCOVERY DESCRIPTION (10) Surveillance test
 ACTIVITY CONTENT (11) RELEASED OF RELEASE (12) Z (13) 33
 AMOUNT OF ACTIVITY (14) NA
 LOCATION OF RELEASE (15) NA
 PERSONNEL EXPOSURES (16) NUMBER (17) 0 (18) 0 (19) 0 (20) 7 (21) Z (22) 38
 DESCRIPTION (23) NA
 PERSONNEL INJURIES (24) NUMBER (25) 0 (26) 0 (27) 0 (28) 40
 DESCRIPTION (29) NA
 LOSS OF OR DAMAGE TO FACILITY (30) TYPE (31) Z (32) 42
 DESCRIPTION (33) NA
 PUBLICITY (34) ISSUED (35) N (36) 44
 DESCRIPTION (37) 8112140179 811204
 PDR ADOCK 05000296
 S PDR
 NRC USE ONLY

NAME OF PREPARED

Gene Holder

PHONE

(205)⁶⁸ 729-6134⁶⁹

LER SUPPLEMENTAL INFORMATION

BFRO-50- 296 / 81066 Technical Specification Involved 3.8.B-6 and 3.8.B-8

Reported Under Technical Specification 6.7.2.b(2) *Date due NRC: 12/7/81

Date of Occurrence 11/7/81 Time of Occurrence 1800 Unit 3

Identification and Description of Occurrence:

CAM 3-RM-90-250 failed to pass on gas and particulate channels.

Conditions Prior to Occurrence:

Unit 1 at 99%.

Unit 2 at 99%.

Unit 3 in refueling outage.

Action specified in the Technical Specification Surveillance Requirements met due to inoperable equipment. Describe.

None

Apparent Cause of Occurrence:

Component failure

Analysis of Occurrence:

There was no danger to the health or safety of the public, no release of activity, no damage to the plant or equipment, and no resulting significant chain of events.

Corrective Action:

The NMC model CRA-14BF gas count rate card was replaced and the continuous air monitor was functionally tested and returned to service. This is considered a random failure and no further recurrence control is required.

Failure Data:

None

Retention: Period - Lifetime; Responsibility - Document Control Supervisor

*Revision:

ALL