

LICENSEE EVENT REPORT

CONTROL BLOCK: _____ (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

0 1 | N | J | S | G | S | 2 | 2 | 0 | 0 | - | 0 | 0 | 0 | 0 | 0 | - | 0 | 0 | 3 | 4 | 1 | 1 | 1 | 1 | 4 | 5

CON'T
0 1 | REPORT SOURCE | L | 6 | 0 | 5 | 0 | 0 | 0 | 3 | 1 | 1 | 7 | 1 | 2 | 1 | 2 | 8 | 1 | 1 | 8 | 0 | 1 | 0 | 8 | 8 | 2 | 9

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)
0 2 | On December 12, 1981 the instrument setpoints on Radiation Monitor 2R12A failed high.
0 3 | At 0717 hours 2R12A was declared inoperable and Action Statement 3.3.3.1 table 3.3-6
0 4 | was entered.
0 5 |
0 6 |
0 7 |
0 8 |

0 9 | SYSTEM CODE | B | A | 11 | CAUSE CODE | X | 12 | CAUSE SUBCODE | Z | 13 | COMPONENT CODE | I | N | S | T | R | U | 14 | COMP. SUBCODE | E | 15 | VALVE SUBCODE | Z | 16
17 | LER/RO REPORT NUMBER | 8 | 1 | 21 | 22 | SEQUENTIAL REPORT NO. | 1 | 2 | 6 | 24 | 26 | OCCURRENCE CODE | 0 | 3 | 28 | 29 | REPORT TYPE | L | 30 | 31 | REVISION NO. | 0 | 32
ACTION TAKEN | E | 18 | 33 | FUTURE ACTION | Z | 19 | 34 | EFFECT ON PLANT | Z | 20 | 35 | SHUTDOWN METHOD | Z | 21 | 36 | HOURS | 0 | 0 | 0 | 0 | 22 | 37 | ATTACHMENT SUBMITTED | Y | 23 | 41 | NPRD-4 FORM SUB. | Y | 24 | 42 | PRIME COMP. SUPPLIER | L | 25 | 43 | COMPONENT MANUFACTURER | V | 1 | 1 | 5 | 26 | 44 | 47

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)
1 0 | The 2R12A setpoints failed high from unknown causes. Subsequent testing has failed
1 1 | to reveal a cause. The monitor was removed and bench-tested satisfactorily. The bat-
1 2 | teries were tested, and were fully charged, but were replaced as a precautionary
1 3 | measure. The monitor was reinstalled, the setpoints reinstated, and it was tested
1 4 | satisfactorily. At 1120 hours 2R12A was declared operable and Action Statement 3.3.3.1
was terminated. (81-101)

1 5 | FACILITY STATUS | E | 28 | 7 | 8 | 9 | % POWER | 0 | 9 | 8 | 29 | 10 | 12 | OTHER STATUS | N/A | 30 | 13 | METHOD OF DISCOVERY | A | 31 | 44 | 45 | 46 | DISCOVERY DESCRIPTION | Setpoints Failed High | 32 | 48

1 6 | ACTIVITY RELEASED | Z | 33 | 7 | 8 | 9 | CONTENT OF RELEASE | Z | 34 | 10 | 11 | AMOUNT OF ACTIVITY | NA | 35 | 44 | LOCATION OF RELEASE | NA | 36 | 45 | 48

1 7 | PERSONNEL EXPOSURES | 0 | 0 | 0 | 37 | 7 | 8 | 9 | TYPE | Z | 38 | 11 | 12 | DESCRIPTION | NA | 39 | 13 | 48

1 8 | PERSONNEL INJURIES | 0 | 0 | 0 | 40 | 7 | 8 | 9 | DESCRIPTION | NA | 41 | 11 | 12 | 48

1 9 | LOSS OF OR DAMAGE TO FACILITY | Z | 42 | 7 | 8 | 9 | TYPE | NA | 43 | 10 | 48

2 0 | PUBLICITY ISSUED | N | 44 | 7 | 8 | 9 | DESCRIPTION | NA | 45 | 10 | 48

8201250411 820108
PDR AD0CK 05000311
S PDR

NRC USE ONLY

NAME OF PREPARER Frank Dickey PHONE: 609-935-0998