

20 November 1981

Mr. Glen D. Brown, Chief
Technical Inspection Branch
U. S. Nuclear Regulatory Commission, Region IV
611 Ryan Plaza Drive, Suite 1000
Arlington, Texas 76011

Dear Sir:

Subject: NRC License No. 17-16380-01
NRC Inspection, Docket 30-10924

Reference: Your letter dated 3 November 1981 with
attached Appendix of NRC Inspection
conducted on 1 October 1981 (Docket:
30-10924; License: 17-16380-01).

Calibration of the pocket dosimeters as per Reference
Appendix, paragraph 1, has been accomplished as required
by 10 CFR 34.33(c), in accordance with calibration in-
structions provided by Victoreen Instrument Division.
All dosimeters performed within acceptable tolerances.
Future calibration will be accomplished on a quarterly
basis.

The annual reports noted in Reference Appendix, para-
graphs 2 and 3, were submitted to the Director of Manage-
ment and Program Analysis on 17 November 1981. Future
reports will be submitted on a timely basis as required
by 10 CFR 20.407(a).

I certify that all information contained in this letter,
including any supplements attached thereto, is true and
correct to the best of my knowledge and belief.

November 20, 1981
Date

Lucas Johnson
Signature

Madge Lombardini
Notary

GENERAL MANAGER
Title

8202010339 820118
NMS LIC30
17-16380-01 PDR