

WYANDOTTE
General Hospital

January 4, 1982

D. J. Sreniawski, Chief
Material Radiation Protection
Section 2
U. S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137


RE: License # 21-12930-01

Dear Mr. Sreniawski:

Attached please find our responses to your notice of violation resulting from the inspection conducted at Wyandotte General Hospital on November 19, 1981. This response is submitted to your office according to the provisions of 10 CFR 2.201.

We affirm the response as submitted here.

Sincerely,



Bruce L. Nelson
Vice President for Medical Services

BLN:mlw

cc: T. McDonald, M.D., Director, Nuclear Medicine
R. J. Nadolny, Executive Vice President
R. Patel, Health Physicist

Encs.: 2

8202020458 820127
NMS LIC30
21-12930-01 PDR

JAN 11 1982

Violations and corrective actions:

1. Quarterly linearity checks:

a. On November 24, 1981, linearity checks were made according to the procedure described in Appendix D, section 2, and was found to be within 5%.

b. Corrective action for future:

The instrument has been tagged with a visible label to indicate the day linearity is to be performed. The label also contains information such as results and comments (see attached example).

2. Survey instrument: Annual calibration

a. Victoreen model 444, serial #389 has been scheduled to be calibrated during the month of January 1982. As soon as the loaner unit is received this unit will be sent for calibration.

b. Corrective action for future:

All the survey instruments have been labeled with a due date for calibration. Calibration has been scheduled to be performed in January and July of every year, thus any further non compliance can be very easily avoided.

3. Leak test:

a. Leak test on cesium-137 sealed source used for dose calibrator was performed on November 19, 1981, the day of inspection, and was found in good order.

b. Corrective action for future:

Routine schedules have been prepared for leak test to accomodate bi-annual leak test for sealed sources.

Enclosure 1

Wyandotte General Hospital

1-4-82

LINEARITY CHECKS

Due Dates	Results	Comments
Jan 11/82		
Apr 12/82		
Jul 19/82		
Oct 18/82		

Enclosure 2

Wyandotte General Hospital
1-4-82