

TEPA

FEB 2 1982

Docket No. 030-01251

Waterbury Hospital Health Center  
ATTN: George W. Curley, Jr.  
Associate Administrator  
64 Robbins Street  
Waterbury, Connecticut 06708



Gentlemen:

Subject: Inspection 81-01

This refers to the routine safety inspection conducted by Mr. C. F. Stearns of this office on December 2, 1981 of activities authorized by NRC License No. 06-02406-01 and to the discussions of our findings held by Mr. Stearns with yourself at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspector, and observations by the inspector.

Our inspector also verified the steps you have taken to correct the items of noncompliance brought to your attention in the enclosure to our letter dated November 25, 1980. We have no further questions regarding your action at this time.

Within the scope of this inspection, no items of noncompliance were observed.

From the discussions at the conclusion of the inspection, it is our understanding that you will make measurements in the adjacent rooms when patients are actually being treated with cesium-137 sealed sources, to ensure that a radiation level of 3 millirem per hour is not exceeded. We will review this matter during our next inspection of your licensed program.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter will be placed in the Public Document Room.

RI:DE&TI  
Stearns/cgl

1/29/82

RI:DE&TI  
Kinneman

2/1/82

OFFICIAL RECORD COPY

8202090146 820202  
NMS LIC30  
06-02406-01 PDR

RIDS: 1607  
5  
1/1

FEB 2 1982

No reply to this letter is required. Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By:

JOHN D. KINNEMAN

Thomas T. Martin, Director  
Division of Engineering and Technical  
Inspection

cc:

Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
State of Connecticut

bcc:

Region I Docket Room (with concurrences)  
Chief, Operational Support Section

OFFICIAL RECORD COPY