

MAR 16 1982

St. Anthony's Hospital
ATTN: Sister Mary Theotima
President
St. Anthony's Way
Alton, Illinois 62002

License No. 12-12189-01

Dear Sister:

Thank you for your letter dated March 1, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated February 3, 1982.

This letter also refers to the telephone conversations on March 13 and 15, 1982, between Mr. Ronald D. Edwards, your consultant, and Mr. W. P. Reichhold of my staff concerning your corrective actions. It is our understanding that you will increase your sample counting time to assure adequate sensitivity to detect low levels of contamination. The procedures for counting wipe test samples will be in effect by April 5, 1982.

If our understanding of the above is incorrect, please contact this office.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

William J. Alden for

D. G. Wiedeman, Acting Chief
Materials Radiation Protection
Section 1

cc w/ltr dated March 1, 1982:
Radiation Protection Consultants, Ltd.
ATTN: Mr. Ronald D. Edwards
Health Physicist
DMB/Document Control Desk (RIDS)

IE HQ FILE COPY

OFFICE	RIII	RIII					
SURNAME	Reichhold:rr	Wiedeman					
DATE	3/10	8203230660	820316				
	NMS LIC30						
	12-12189-01	PDR					
NRC FORM 318 (Rev. 11-79) NRC 0240							

OFFICIAL RECORD COPY

DMB:TE07
5/11