

NOTICE OF VIOLATION

Roche Biomedical Labs., Inc.
Kansas City, Missouri, 64130

Docket No. 030-21056
License No. 24-23424-01

As a result of the inspection conducted on March 27, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

License Condition 14 requires that all licensed material be possessed and used in accordance with statements, representations, and procedures contained in the referenced application dated March 15, 1989.

The application dated March 15, 1989, states that swipe tests will be performed weekly of any area where radioactive materials are used or stored that has the potential for contamination of work surfaces and floors.

Contrary to the above, swipe tests were not performed weekly to check the work surfaces and floors for contamination. Specifically, weekly swipe tests were not conducted from February 19, 1990 until the date of inspection March 27, 1990 in laboratory areas where radioactive materials are routinely used and stored.

This is a repeat violation.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

April 12, 1990
Dated

Roy J. Caniano
Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2