

NOTICE OF VIOLATION

Holland Community Hospital
Holland, MI 49423-4999

Docket No. 030-13801
License No. 21-18502-01

As a result of the inspection conducted on June 5, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

10 CFR 35.59(g) requires that a quarterly inventory be conducted of sealed (calibration) sources in possession of the licensee.


Contrary to the above, the required inventory of calibration sources was not conducted from October 28, 1986 to February 9, 1988.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

June 19, 1990


Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2

9006260528 900619
REG3 LIC30
21-18502-01 PDC