

Appendix

NOTICE OF VIOLATION

Medical College of Ohio at Toledo

License No. 34-13011-04

As a result of the inspection conducted on April 19, 20 and 23, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. 10 CFR 20.201(b) requires each licensee to make or cause to be made such surveys (evaluations) as may be necessary for him to comply with the regulations in Part 20.

Contrary to this requirement, you failed to survey the radiation levels in unrestricted areas adjacent to an implant patient room on April 12, 1982, to determine compliance with 10 CFR 20.105(b).

This is a Severity Level IV violation (Supplement IV).

2. 10 CFR 35.14(b)(5)(v) requires a quarterly physical inventory to account for all sealed sources and devices received and possessed.

Contrary to this requirement, your physical inventory of such sources on October 19, 1981 did not include ten cesium-137 Micrad sources, approximately 26 millicuries each, that were in your possession at that time.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated JUL 15 1982

*Original*

*S/ly James G. Keppler*  
James G. Keppler  
Regional Administrator