

NOTICE OF VIOLATION

Henry R. Silverman, III, M.D.  
Toledo, OH

License No. 34-20317-01  
Docket No. 030-18243

As a result of the followup conducted from August 3 to October 2, 1990, and in accordance with 10 CFR Part 2, Appendix C - General Statement of Policy and Procedure for NRC Enforcement Actions (1989), the following violation was identified:

License Condition No. 10 authorizes and limits the possession and use of licensed materials at Doctor's Building, 3939 Monroe Street, Toledo, Ohio.

Contrary to the above, since July 20, 1990, the licensee has possessed a nominal 200 microcurie cesium-137 sealed source at 142 East Maumee Street, Adrian, Michigan, a location not authorized by the NRC license.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective actions that have been taken and the results achieved; (2) the corrective actions that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

October 12, 1990  
Dated

John A. Grobe  
John A. Grobe, Chief  
Nuclear Materials Safety Branch