

NOTICE OF VIOLATION

Leeco Diagnostics, Inc.
Southfield, MI

License No. 21-18294-01
Docket No. 030-14802

As a result of the inspection conducted on September 26, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violations were identified:

1. License Condition No. 14 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced documents.

The referenced application dated August 31, 1978 requires that film badges and TLD ring badges be worn by employees at all times when working with radioactive material.

Contrary to the above, on September 26, 1990, a licensee employee was observed handling millicurie quantities of iodine-125 during an iodination process, while not wearing his assigned TLD ring badge.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 14 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced documents.


The referenced letter dated March 26, 1986, requires that all iodinations take place in large flow hoods with continuous flow indicators and activated charcoal filters.

Contrary to the above, since July 1989, iodinations have been routinely performed in hoods not equipped with activated charcoal filters.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

10-26-90
Date



John A. Grobe, Chief
Nuclear Materials Safety
Branch

9011130210 901026
REGS LIC30 FDC
21-18294-01