

NOTICE OF VIOLATION

Standish Community Hospital
P.O. Box 579
Standish, MI 48658

License No. 21-21265-01
Doc: et No. 030-20121

As a result of the inspection conducted on October 9, 1990, and in accordance with 10 CFR Part 2, Appendix C - General Statement of Policy and Procedure for NRC Enforcement Actions (1989), the following violations were identified:

1. License Condition No. 11 states that a specified individual is the Radiation Safety Officer.

Contrary to the above, from June 4, 1990 until July 16, 1990, the specified individual did not serve as Radiation Safety Officer.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 35.22(a)(2) requires that the Radiation Safety Committee meet at least quarterly.


Contrary to the above, during the second quarter of 1990, the licensee's Radiation Safety Committee failed to meet.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective actions that have been taken and the results achieved; (2) the corrective actions that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

NOV 05 1990

Dated _____


John A. Grobe, Chief
Nuclear Materials Safety Branch

9011190103 901105
REG3 LIC30
21-21265-01 PDC