

GLENDIVE COMMUNITY HOSPITAL
COMMUNITY OWNED-NON-PROFIT

PROSPECT AND AMES
GLENDIVE, MONTANA 59330
TELEPHONE 406-365-3306

July 12, 1982


Glen D. Brown, Chief
Technical Inspection Branch
U.S. Nuclear Regulatory Commission
611 Ryan Plaza Drive, Suite 1000
Arlington, Texas 76011

Dear Mr. Brown;

The deficiencies cited in your letter of December 22, 1981, have been discussed at our isotope meeting held at the end of March. The isotope committee has accepted the responsibility to monitor annual calibration of the dose calibrator. The Medical Staff will review the meetings of the isotope committee to assure their compliance with a regular quarterly meeting schedule.

This should answer your remaining questions concerning these deficiencies.

Sincerely,



Donald H. Botkin
Administrator

"I certify that all information contained in this letter, including any supplements attached thereto, is true and correct to the best of my knowledge and belief."

July 12, 1982
DATE

Donald H. Botkin
ADMINISTRATOR

Subscribed and sworn to before me this 12th day of July, 1982.

Scott Anderson
Notary Public
Notary Public for the State of Montana
County of Dawson
Residing at Glendive, Montana. My
commission expires in 1983.